

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN012820\
 Data File : VN059896.D
 Acq On : 28 Jan 2020 14:42
 Operator : JC/MD
 Sample : L1267-03 500X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 40329

Quant Time: Jan 28 22:46:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	208561	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	309551	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	268157	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	122419	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	106730	46.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.16%	
35) Dibromofluoromethane	7.56	113	92582	50.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.30%	
50) Toluene-d8	10.08	98	359982	50.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.66%	
62) 4-Bromofluorobenzene	12.40	95	120475	46.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.44%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.06	83	2457	0.739	ug/l #	84
40) Benzene	8.02	78	3704	0.461	ug/l	97
52) Toluene	10.14	92	16311	3.212	ug/l	96
67) Ethyl Benzene	11.50	91	70874	7.407	ug/l	98
68) m/p-Xylenes	11.61	106	116100	32.123	ug/l	93
69) o-Xylene	11.94	106	70118	19.816	ug/l	95
73) Isopropylbenzene	12.24	105	37900	4.221	ug/l	100
78) n-propylbenzene	12.59	91	136886	13.584	ug/l	97
80) 1,3,5-Trimethylbenzene	12.73	105	187021	24.769	ug/l	100
84) 1,2,4-Trimethylbenzene	13.03	105	806444	107.959	ug/l	98
85) sec-Butylbenzene	13.17	105	80157	9.342	ug/l	87
86) p-Isopropyltoluene	13.28	119	53533	6.743	ug/l	97
89) n-Butylbenzene	13.61	91	112133	16.566	ug/l #	78
95) Naphthalene	15.13	128	243977	36.561	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN012820\
Data File : VN059896.D
Acq On : 28 Jan 2020 14:42
Operator : JC/MD
Sample : L1267-03 500X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
40329

Quant Time: Jan 28 22:46:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

