

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065645.D
 Acq On : 28 Jan 2021 13:16
 Operator : JC/MD
 Sample : M1273-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-110

Quant Time: Jan 29 01:22:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.624	168	212773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	371187	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	321255	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	120763	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	135840	48.53	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.06%		
35) Dibromofluoromethane	7.550	113	108708	49.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.84%		
50) Toluene-d8	10.058	98	463140	53.56	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.12%		
62) 4-Bromofluorobenzene	12.373	95	152884	49.10	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.20%		
Target Compounds					Qvalue	
13) Acrolein	3.573	56	281	18.66	ug/l #	28
16) Acetone	3.782	43	4405	4.04	ug/l	88
17) Carbon Disulfide	4.007	76	1668	0.20	ug/l #	74
25) 2-Butanone	6.814	43	1342	0.79	ug/l #	46
28) Bromochloromethane	7.332	49	1371	0.62	ug/l #	9
30) Chloroform	7.328	83	19804	4.11	ug/l	92
31) Cyclohexane	7.624	56	4420	0.84	ug/l #	1
36) 1,1-Dichloropropene	7.624	75	17391	4.66	ug/l #	47
38) Carbon Tetrachloride	7.624	117	20233	6.58	ug/l #	16
39) Methylcyclohexane	9.042	83	1176	0.26	ug/l	98
47) Bromodichloromethane	9.370	83	1290	0.35	ug/l #	92
55) 1,1,2-Trichloroethane	10.499	97	1121	0.43	ug/l #	15
59) 2-Hexanone	10.730	43	529	0.22	ug/l #	23
67) Ethyl Benzene	11.483	91	5581	0.44	ug/l	97
68) m/p-Xylenes	11.595	106	1020	0.22	ug/l	99
69) o-Xylene	11.917	106	11902	2.69	ug/l	92
71) Bromoform	12.373	173	657	0.40	ug/l #	1
73) Isopropylbenzene	12.219	105	9388	0.89	ug/l	95
76) 1,2,3-Trichloropropane	12.370	75	69852	22.12	ug/l #	100
78) n-propylbenzene	12.563	91	11531	0.93	ug/l	95
80) 1,3,5-Trimethylbenzene	12.704	105	24886	2.76	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.370	75	69852	56.66	ug/l #	12
82) 4-Chlorotoluene	12.704	91	2412	0.32	ug/l #	40
84) 1,2,4-Trimethylbenzene	13.010	105	40777	4.54	ug/l	98
85) sec-Butylbenzene	13.142	105	7111	0.70	ug/l	99
86) p-Isopropyltoluene	13.261	119	7663	0.86	ug/l	97
89) n-Butylbenzene	13.553	91	2365	0.29	ug/l #	1
90) Hexachloroethane	13.859	117	2917	1.86	ug/l #	12
93) 1,2,4-Trichlorobenzene	14.881	180	63	3.39	ug/l #	1
95) Naphthalene	15.100	128	8239	4.43	ug/l #	86
96) 1,2,3-Trichlorobenzene	15.277	180	207	2.93	ug/l #	35

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065645.D
Acq On : 28 Jan 2021 13:16
Operator : JC/MD
Sample : M1273-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP-110

Quant Time: Jan 29 01:22:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampled :
WP-110

Tgt Ion:168 Resp: 212773
Ion Ratio Lower Upper
168 100
99 59.7 42.6 63.8



#13
Acrolein
Concen: 18.66 ug/l
RT: 3.573 min Scan# 570
Delta R.T. 0.004 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 56 Resp: 281
Ion Ratio Lower Upper
56 100
55 10.7 56.0 84.0#









#16
Acetone
Concen: 4.04 ug/l
RT: 3.782 min Scan# 635
Delta R.T. -0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampled :
WP-110


Tgt Ion: 43 Resp: 4405
Ion Ratio Lower Upper
43 100
58 40.4 27.0 40.6



#17
Carbon Disulfide
Concen: 0.20 ug/l
RT: 4.007 min Scan# 705
Delta R.T. -0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 76 Resp: 1668
Ion Ratio Lower Upper
76 100
78 0.0 7.8 11.6#









#25
2-Butanone
Concen: 0.79 ug/l
RT: 6.814 min Scan# 1578
Delta R.T. 0.023 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampled :
WP-110


Tgt Ion: 43 Resp: 1342
Ion Ratio Lower Upper
43 100
72 0.0 23.1 34.7#



#28
Bromochloromethane
Concen: 0.62 ug/l
RT: 7.332 min Scan# 1739
Delta R.T. 0.184 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 49 Resp: 1371
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.6
130 0.0 67.8 101.6#









#30
Chloroform
Concen: 4.11 ug/l
RT: 7.328 min Scan# 1738
Delta R.T. 0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion: 83 Resp: 19804
Ion Ratio Lower Upper
83 100
85 59.6 52.6 78.8



#31
Cyclohexane
Concen: 0.84 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. 0.013 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 56 Resp: 4420
Ion Ratio Lower Upper
56 100
69 220.2 26.6 39.8#
84 199.6 70.4 105.6#









#33
1,2-Dichloroethane-d4
Concen: 48.53 ug/l
RT: 7.988 min Scan# 1943
Delta R.T. 0.001 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion: 65 Resp: 135840
Ion Ratio Lower Upper
65 100
67 56.8 0.0 107.2



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.547 min Scan# 2117
Delta R.T. -0.003 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 114 Resp: 371187
Ion Ratio Lower Upper
114 100
63 22.4 0.0 38.6
88 16.6 0.0 30.6





#35

Dibromofluoromethane

Concen: 49.42 ug/l

RT: 7.550 min Scan# 1807

Delta R.T. 0.003 min

Lab File: VN065645.D

Acq: 28 Jan 2021 13:16





Tgt Ion:113 Resp: 108708

Ion Ratio Lower Upper

113 100

111 103.8 82.2 123.2

192 17.0 15.4 23.0



#36

1,1-Dichloropropene

Concen: 4.66 ug/l

RT: 7.624 min Scan# 1830

Delta R.T. -0.125 min

Lab File: VN065645.D

Acq: 28 Jan 2021 13:16




Tgt Ion: 75 Resp: 17391


Ion Ratio Lower Upper

75 100

110 7.3 18.2 54.6#


77 0.0 24.9 37.3#








#38
Carbon Tetrachloride
Concen: 6.58 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. -0.106 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Instrument :
MSVOA_N
ClientSampleId :
WP-110







Tgt Ion: 117 Resp: 20233
Ion Ratio Lower Upper
117 100
119 5.0 76.0 114.0#
121 0.0 24.5 36.7#



#39
Methylcyclohexane
Concen: 0.26 ug/l
RT: 9.042 min Scan# 2271
Delta R.T. 0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 83 Resp: 1176
Ion Ratio Lower Upper
83 100
55 74.7 58.9 88.3
98 42.7 35.9 53.9









#47
Bromodichloromethane
Concen: 0.35 ug/l
RT: 9.370 min Scan# 2373
Delta R.T. 0.003 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampled :
WP-110


Tgt Ion: 83 Resp: 1290
Ion Ratio Lower Upper
83 100
85 58.8 51.0 76.4
127 0.0 7.0 10.6#



#50
Toluene-d8
Concen: 53.56 ug/l
RT: 10.058 min Scan# 2587
Delta R.T. 0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 98 Resp: 463140
Ion Ratio Lower Upper
98 100
100 64.1 50.5 75.7









#55
1,1,2-Trichloroethane
Concen: 0.43 ug/l
RT: 10.499 min Scan# 2724
Delta R.T. -0.032 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion: 97	Resp: 1121
Ion Ratio	Lower Upper
97 100	
83 0.0	67.7 101.5#
85 0.0	44.5 66.7#
99 0.0	49.4 74.0#



#59
2-Hexanone
Concen: 0.22 ug/l
RT: 10.730 min Scan# 2796
Delta R.T. 0.010 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 43	Resp: 529
Ion Ratio	Lower Upper
43 100	
58 0.0	28.1 84.3#









#62
4-Bromofluorobenzene
Concen: 49.10 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. 0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion: 95 Resp: 152884
Ion Ratio Lower Upper
95 100
174 67.0 0.0 169.0
176 65.1 0.0 162.4



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.376 min Scan# 2997
Delta R.T. 0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 117 Resp: 321255
Ion Ratio Lower Upper
117 100
82 58.2 43.7 65.5
119 31.3 26.1 39.1









#67
Ethyl Benzene
Concen: 0.44 ug/l
RT: 11.483 min Scan# 3030
Delta R.T. 0.004 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion: 91 Resp: 5581
Ion Ratio Lower Upper
91 100
106 30.4 25.5 38.3



#68
m/p-Xylenes
Concen: 0.22 ug/l
RT: 11.595 min Scan# 3065
Delta R.T. 0.003 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 106 Resp: 1020
Ion Ratio Lower Upper
106 100
91 203.2 161.4 242.0









#69
o-Xylene
Concen: 2.69 ug/l
RT: 11.917 min Scan# 3165
Delta R.T. 0.001 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampled :
WP-110


Tgt Ion:106 Resp: 11902
Ion Ratio Lower Upper
106 100
91 225.5 106.8 320.4



#71
Bromoform
Concen: 0.40 ug/l
RT: 12.373 min Scan# 3307
Delta R.T. 0.277 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion:173 Resp: 657
Ion Ratio Lower Upper
173 100
175 942.5 23.8 71.2#
254 0.0 0.0 0.0









#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.312 min Scan# 3599
Delta R.T. 0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion:152 Resp: 120763
Ion Ratio Lower Upper
152 100
115 63.4 28.5 85.6
150 158.1 0.0 347.6



#73
Isopropylbenzene
Concen: 0.89 ug/l
RT: 12.219 min Scan# 3259
Delta R.T. -0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion:105 Resp: 9388
Ion Ratio Lower Upper
105 100
120 24.0 13.3 39.9









#76
1,2,3-Trichloropropane
Concen: 22.12 ug/l
RT: 12.370 min Scan# 3306
Delta R.T. -0.154 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion: 75 Resp: 69852
Ion Ratio Lower Upper
75 100
77 1.6 0.0 0.0#



#78
n-propylbenzene
Concen: 0.93 ug/l
RT: 12.563 min Scan# 3366
Delta R.T. -0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 91 Resp: 11531
Ion Ratio Lower Upper
91 100
120 21.1 11.8 35.3









#80
1,3,5-Trimethylbenzene
Concen: 2.76 ug/l
RT: 12.704 min Scan# 3410
Delta R.T. 0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion:105 Resp: 24886
Ion Ratio Lower Upper
105 100
120 46.8 24.3 72.8



#81
trans-1,4-Dichloro-2-butene
Concen: 56.66 ug/l
RT: 12.370 min Scan# 3306
Delta R.T. 0.100 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 75 Resp: 69852
Ion Ratio Lower Upper
75 100
53 0.0 74.6 111.8#
89 0.0 35.8 53.8#









#82
4-Chlorotoluene
Concen: 0.32 ug/l
RT: 12.704 min Scan# 3410
Delta R.T. -0.039 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion: 91 Resp: 2412
Ion Ratio Lower Upper
91 100
126 0.0 17.2 51.6#



#84
1,2,4-Trimethylbenzene
Concen: 4.54 ug/l
RT: 13.010 min Scan# 3505
Delta R.T. -0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 105 Resp: 40777
Ion Ratio Lower Upper
105 100
120 45.7 22.4 67.0









#85
 sec-Butylbenzene
 Concen: 0.70 ug/l
 RT: 13.142 min Scan# 3546
 Delta R.T. 0.001 min
 Lab File: VN065645.D
 Acq: 28 Jan 2021 13:16




Instrument :
 MSVOA_N
 ClientSampled :
 WP-110


Tgt Ion:105 Resp: 7111
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.0 30.0



#86
 p-Isopropyltoluene
 Concen: 0.86 ug/l
 RT: 13.261 min Scan# 3583
 Delta R.T. 0.004 min
 Lab File: VN065645.D
 Acq: 28 Jan 2021 13:16

Tgt Ion:119 Resp: 7663
 Ion Ratio Lower Upper
 119 100
 134 24.5 13.3 39.8
 91 22.4 11.8 35.3









#89
n-Butylbenzene
Concen: 0.29 ug/l
RT: 13.553 min Scan# 3674
Delta R.T. -0.032 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampleId :
WP-110


Tgt Ion: 91 Resp: 2365
Ion Ratio Lower Upper
91 100
92 22.2 26.1 78.3#
134 509.1 13.0 38.9#



#90
Hexachloroethane
Concen: 1.86 ug/l
RT: 13.859 min Scan# 3769
Delta R.T. 0.017 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16

Tgt Ion: 117 Resp: 2917
Ion Ratio Lower Upper
117 100
201 0.0 37.8 113.3#









#93
1,2,4-Trichlorobenzene
Concen: 3.39 ug/l
RT: 14.881 min Scan# 4087
Delta R.T. 0.003 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16




Instrument :
MSVOA_N
ClientSampled :
WP-110


Tgt Ion:180 Resp: 63
Ion Ratio Lower Upper
180 100
182 98.4 47.3 142.0
145 6576.2 15.6 46.8#



#95
Naphthalene
Concen: 4.43 ug/l
RT: 15.100 min Scan# 4155
Delta R.T. -0.000 min
Lab File: VN065645.D
Acq: 28 Jan 2021 13:16


Tgt Ion:128 Resp: 8239
Ion Ratio Lower Upper
128 100
127 16.0 10.2 15.2#
129 19.0 8.6 13.0#





#96
 1,2,3-Trichlorobenzene
 Concen: 2.93 ug/l
 RT: 15.277 min Scan# 4210
 Delta R.T. -0.003 min
 Lab File: VN065645.D
 Acq: 28 Jan 2021 13:16

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-110



Tgt Ion:180 Resp: 207
 Ion Ratio Lower Upper
 180 100
 182 29.0 47.4 142.3#
 145 0.0 16.5 49.5#

