

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065644.D
 Acq On : 28 Jan 2021 12:51
 Operator : JC/MD
 Sample : M1272-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-325

Quant Time: Jan 29 01:21:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	200367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	347227	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	293450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	103253	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	125144	47.47	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.94%
35) Dibromofluoromethane	7.550	113	100830	49.00	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.00%
50) Toluene-d8	10.058	98	429008	53.04	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.08%
62) 4-Bromofluorobenzene	12.373	95	135932	46.67	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.34%
Target Compounds						
					Qvalue	
16) Acetone	3.785	43	4925	4.80	ug/l	100
30) Chloroform	7.328	83	4933	1.09	ug/l	98
95) Naphthalene	15.100	128	9027	4.71	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065644.D
Acq On : 28 Jan 2021 12:51
Operator : JC/MD
Sample : M1272-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP-325

Quant Time: Jan 29 01:21:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

