

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065645.D
 Acq On : 28 Jan 2021 13:16
 Operator : JC/MD
 Sample : M1273-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-110

Manual Integrations
 APPROVED

MMDadoda
 1/29/2021 12:00:13 PM

Quant Time: Jan 29 01:22:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	212773	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	371187	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	321255	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	120763	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	135840	48.53	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.06%		
35) Dibromofluoromethane	7.550	113	108708	49.42	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.84%		
50) Toluene-d8	10.058	98	463140	53.56	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.12%		
62) 4-Bromofluorobenzene	12.373	95	152884	49.10	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.20%		
Target Compounds						
					Qvalue	
16) Acetone	3.782	43	4405	4.04	ug/l	88
30) Chloroform	7.328	83	19804	4.11	ug/l	92
39) Methylcyclohexane	9.042	83	1176	0.26	ug/l	98
67) Ethyl Benzene	11.483	91	5581	0.44	ug/l	97
68) m/p-Xylenes	11.595	106	1020	0.22	ug/l	99
69) o-Xylene	11.917	106	11902	2.69	ug/l	92
73) Isopropylbenzene	12.219	105	9388	0.89	ug/l	95
78) n-propylbenzene	12.563	91	11531	0.93	ug/l	95
80) 1,3,5-Trimethylbenzene	12.704	105	24886	2.76	ug/l	97
84) 1,2,4-Trimethylbenzene	13.010	105	40777	4.54	ug/l	98
85) sec-Butylbenzene	13.142	105	7111	0.70	ug/l	99
86) p-Isopropyltoluene	13.261	119	7663	0.86	ug/l	97
89) n-Butylbenzene	13.585	91	3219m	0.40	ug/l	
95) Naphthalene	15.100	128	8239	4.43	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065645.D
Acq On : 28 Jan 2021 13:16
Operator : JC/MD
Sample : M1273-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP-110

Manual Integrations
APPROVED

MMDadoda
1/29/2021 12:00:13 PM

Quant Time: Jan 29 01:22:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

