

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065646.D
 Acq On : 28 Jan 2021 13:49
 Operator : JC/MD
 Sample : PB134373TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134373TB

Quant Time: Jan 29 01:22:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.621	168	222392	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	387489	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	337349	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	122828	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	139474	47.67	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.34%
35) Dibromofluoromethane	7.547	113	111018	48.35	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.70%
50) Toluene-d8	10.055	98	480020	53.18	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.36%
62) 4-Bromofluorobenzene	12.370	95	162103	49.87	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.74%
Target Compounds						
16) Acetone	3.782	43	6538	5.74	ug/l	92
20) Methylene Chloride	4.499	84	47896	15.19	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065646.D
Acq On : 28 Jan 2021 13:49
Operator : JC/MD
Sample : PB134373TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134373TB

Quant Time: Jan 29 01:22:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

