

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065650.D
 Acq On : 28 Jan 2021 15:27
 Operator : JC/MD
 Sample : PB134373ZHE#13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134373ZHE#13

Quant Time: Jan 29 01:23:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	204944	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	358124	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	306966	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	111709	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	130491	48.40	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.80%		
35) Dibromofluoromethane	7.547	113	104425	49.21	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	98.42%		
50) Toluene-d8	10.058	98	446006	53.46	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.92%		
62) 4-Bromofluorobenzene	12.373	95	144807	48.21	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.42%		
Target Compounds					Qvalue	
16) Acetone	3.782	43	5945	5.66	ug/l	98
20) Methylene Chloride	4.502	84	38086	13.10	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065650.D
Acq On : 28 Jan 2021 15:27
Operator : JC/MD
Sample : PB134373ZHE#13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134373ZHE#13

Quant Time: Jan 29 01:23:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

