

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065653.D
 Acq On : 28 Jan 2021 16:40
 Operator : JC/MD
 Sample : PB134372ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134372ZHE#01

Quant Time: Jan 29 01:24:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	199529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	348433	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	298651	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	107734	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	126871	48.33	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.66%
35) Dibromofluoromethane	7.550	113	99037	47.96	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.92%
50) Toluene-d8	10.058	98	433767	53.44	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.88%
62) 4-Bromofluorobenzene	12.373	95	140944	48.22	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.44%
Target Compounds						
					Qvalue	
16) Acetone	3.782	43	26719	26.15	ug/l	99
20) Methylene Chloride	4.502	84	40618	14.35	ug/l	94
43) Isopropyl Acetate	8.129	43	18795	3.15	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065653.D
Acq On : 28 Jan 2021 16:40
Operator : JC/MD
Sample : PB134372ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134372ZHE#01

Quant Time: Jan 29 01:24:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

