

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065657.D
 Acq On : 28 Jan 2021 18:18
 Operator : JC/MD
 Sample : PB134372ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134372ZHE#05

Quant Time: Jan 29 01:25:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	210896	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	368701	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	321843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	119066	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	133831	48.24	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.48%		
35) Dibromofluoromethane	7.547	113	106233	48.62	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.24%		
50) Toluene-d8	10.058	98	456363	53.14	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	106.28%		
62) 4-Bromofluorobenzene	12.373	95	154222	49.87	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	99.74%		
Target Compounds						Qvalue
16) Acetone	3.782	43	21032	19.47	ug/l	94
20) Methylene Chloride	4.502	84	12621	4.22	ug/l	97
43) Isopropyl Acetate	8.129	43	20305	3.22	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065657.D
Acq On : 28 Jan 2021 18:18
Operator : JC/MD
Sample : PB134372ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134372ZHE#05

Quant Time: Jan 29 01:25:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

