

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065658.D
 Acq On : 28 Jan 2021 18:43
 Operator : JC/MD
 Sample : PB134372ZHE#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134372ZHE#06

Quant Time: Jan 29 01:25:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	200143	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	348733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	307749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	116617	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	128187	48.68	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.36%
35) Dibromofluoromethane	7.547	113	100120	48.45	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.90%
50) Toluene-d8	10.058	98	434929	53.54	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.08%
62) 4-Bromofluorobenzene	12.373	95	146977	50.25	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.50%
Target Compounds						
16) Acetone	3.782	43	20626	20.12	ug/l	99
18) Methyl Acetate	4.290	43	3363	1.17	ug/l #	66
20) Methylene Chloride	4.502	84	12500	4.40	ug/l	98
43) Isopropyl Acetate	8.129	43	20472	3.43	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065658.D
Acq On : 28 Jan 2021 18:43
Operator : JC/MD
Sample : PB134372ZHE#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134372ZHE#06

Quant Time: Jan 29 01:25:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

