

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
 Data File : VN065663.D
 Acq On : 28 Jan 2021 20:45
 Operator : JC/MD
 Sample : M1255-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CRT-13

Quant Time: Jan 29 01:26:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	211027	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	367482	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	318360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	113587	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	131615	47.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.82%
35) Dibromofluoromethane	7.550	113	105770	48.57	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.14%
50) Toluene-d8	10.058	98	453717	53.00	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.00%
62) 4-Bromofluorobenzene	12.373	95	147647	47.90	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.80%
Target Compounds						
16) Acetone	3.782	43	5694	5.27	ug/l	93
20) Methylene Chloride	4.502	84	155508	51.96	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012821\
Data File : VN065663.D
Acq On : 28 Jan 2021 20:45
Operator : JC/MD
Sample : M1255-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CRT-13

Quant Time: Jan 29 01:26:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

