

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012822\
 Data File : VN070820.D
 Acq On : 28 Jan 2022 11:18
 Operator : JC/MD
 Sample : VN0128WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0128WBL01

Quant Time: Jan 28 23:15:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	768467	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.968	114	1268223	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1157489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	456811	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	564840	51.287	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.580%
35) Dibromofluoromethane	8.021	113	376215	47.823	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.640%
50) Toluene-d8	10.440	98	1563304	47.171	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.340%
62) 4-Bromofluorobenzene	12.731	95	595871	50.457	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012822\
Data File : VN070820.D
Acq On : 28 Jan 2022 11:18
Operator : JC/MD
Sample : VN0128WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0128WBL01

Quant Time: Jan 28 23:15:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

