

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012822\
 Data File : VN070828.D
 Acq On : 28 Jan 2022 14:54
 Operator : JC/MD
 Sample : PB142316ZHE#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142316ZHE#05

Quant Time: Jan 28 23:17:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	944530	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1556403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1461948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	725543	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	642703	47.479	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	94.960%		
35) Dibromofluoromethane	8.024	113	438769	45.448	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	90.900%		
50) Toluene-d8	10.440	98	1957604	48.132	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.260%		
62) 4-Bromofluorobenzene	12.731	95	744578	51.375	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	102.740%		
Target Compounds						Qvalue
20) Methylene Chloride	5.093	84	22673	1.832	ug/l	95
43) Isopropyl Acetate	8.563	43	118142	3.585	ug/l #	84
52) Toluene	10.505	92	40933	1.359	ug/l	93
67) Ethyl Benzene	11.843	91	100272	1.707	ug/l	99
68) m/p-Xylenes	11.950	106	177126	8.291	ug/l	81
69) o-Xylene	12.278	106	28867	1.362	ug/l	96
95) Naphthalene	15.504	128	104227	7.445	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012822\
Data File : VN070828.D
Acq On : 28 Jan 2022 14:54
Operator : JC/MD
Sample : PB142316ZHE#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142316ZHE#05

Quant Time: Jan 28 23:17:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

