

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012822\
 Data File : VN070839.D
 Acq On : 28 Jan 2022 19:34
 Operator : JC/MD
 Sample : N1303-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CC-9

Quant Time: Jan 28 23:19:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	945072	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1579612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1504078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	617055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	650904	48.057	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.120%
35) Dibromofluoromethane	8.024	113	442649	45.176	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.360%
50) Toluene-d8	10.440	98	1975208	47.851	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.700%
62) 4-Bromofluorobenzene	12.731	95	701063	47.662	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.320%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	202692	21.561	ug/l	99
20) Methylene Chloride	5.095	84	13613	1.099	ug/l #	86
67) Ethyl Benzene	11.843	91	118542	1.962	ug/l	96
68) m/p-Xylenes	11.948	106	185965	8.461	ug/l	91
69) o-Xylene	12.278	106	83455	3.826	ug/l	92
80) 1,3,5-Trimethylbenzene	13.058	105	96912	2.053	ug/l	94
84) 1,2,4-Trimethylbenzene	13.369	105	323847	7.137	ug/l	96
95) Naphthalene	15.507	128	129966	8.608	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012822\
Data File : VN070839.D
Acq On : 28 Jan 2022 19:34
Operator : JC/MD
Sample : N1303-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CC-9

Quant Time: Jan 28 23:19:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

