

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012822\
 Data File : VN070841.D
 Acq On : 28 Jan 2022 20:25
 Operator : JC/MD
 Sample : PB142316ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142316ZHE#04

Quant Time: Jan 28 23:19:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	864100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1448885	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1362267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	549765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	599710	48.427	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.860%
35) Dibromofluoromethane	8.021	113	417881	46.496	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.000%
50) Toluene-d8	10.440	98	1821592	48.111	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.220%
62) 4-Bromofluorobenzene	12.731	95	636540	47.180	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.360%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	87259	10.152	ug/l	99
20) Methylene Chloride	5.100	84	21725	1.919	ug/l	89
43) Isopropyl Acetate	8.563	43	116379	3.793	ug/l #	84
95) Naphthalene	15.504	128	18750	5.541	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012822\
Data File : VN070841.D
Acq On : 28 Jan 2022 20:25
Operator : JC/MD
Sample : PB142316ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142316ZHE#04

Quant Time: Jan 28 23:19:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

