

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085532.D
 Acq On : 28 Jan 2025 12:36
 Operator : JC\MD
 Sample : Q1180-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jan 29 01:34:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	325340	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	283619	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	106780	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	156340	53.723	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.440%
35) Dibromofluoromethane	8.165	113	119192	52.809	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.620%
50) Toluene-d8	10.565	98	400005	49.880	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.760%
62) 4-Bromofluorobenzene	12.847	95	117774	42.933	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	85.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085532.D
Acq On : 28 Jan 2025 12:36
Operator : JC\MD
Sample : Q1180-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Jan 29 01:34:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

