

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085533.D
 Acq On : 28 Jan 2025 13:00
 Operator : JC\MD
 Sample : Q1194-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB

Quant Time: Jan 29 01:33:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	169030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	313197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	274340	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	106996	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	149462	54.778	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	109.560%	
35) Dibromofluoromethane	8.165	113	112628	51.835	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.660%	
50) Toluene-d8	10.565	98	381699	49.443	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.880%	
62) 4-Bromofluorobenzene	12.847	95	114691	43.430	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	86.860%	
Target Compounds						
16) Acetone	4.430	43	12983	14.727	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085533.D
Acq On : 28 Jan 2025 13:00
Operator : JC\MD
Sample : Q1194-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB

Quant Time: Jan 29 01:33:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

