

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085534.D
 Acq On : 28 Jan 2025 13:24
 Operator : JC\MD
 Sample : Q1199-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-TB-20250123

Quant Time: Jan 29 00:38:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	179083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	333621	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	293971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	113614	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	160235	55.430	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 110.860%		
35) Dibromofluoromethane	8.165	113	120678	52.140	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.280%		
50) Toluene-d8	10.565	98	407027	49.496	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.000%		
62) 4-Bromofluorobenzene	12.847	95	127940	45.481	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 90.960%		

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085534.D
Acq On : 28 Jan 2025 13:24
Operator : JC\MD
Sample : Q1199-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-TB-20250123

Quant Time: Jan 29 00:38:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

