

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085535.D
 Acq On : 28 Jan 2025 13:49
 Operator : JC\MD
 Sample : Q1199-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-EB-20250124

Quant Time: Jan 29 00:38:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	293573	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	250987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	144816	56.663	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.320%
35) Dibromofluoromethane	8.165	113	109419	53.724	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.440%
50) Toluene-d8	10.565	98	357973	49.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.940%
62) 4-Bromofluorobenzene	12.847	95	107196	43.305	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	86.620%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085535.D
Acq On : 28 Jan 2025 13:49
Operator : JC\MD
Sample : Q1199-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-EB-20250124

Quant Time: Jan 29 00:38:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

