

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085537.D
 Acq On : 28 Jan 2025 14:37
 Operator : JC\MD
 Sample : Q1199-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-260-262

Quant Time: Jan 29 00:39:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	287847	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	90634	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	141633	57.181	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.360%			
35) Dibromofluoromethane	8.165	113	107001	53.582	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.160%			
50) Toluene-d8	10.565	98	346432	48.827	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.660%			
62) 4-Bromofluorobenzene	12.847	95	105662	43.535	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 87.060%			
Target Compounds						
16) Acetone	4.447	43	5039	6.296	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085537.D
Acq On : 28 Jan 2025 14:37
Operator : JC\MD
Sample : Q1199-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-260-262

Quant Time: Jan 29 00:39:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

