

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085538.D
 Acq On : 28 Jan 2025 15:01
 Operator : JC\MD
 Sample : Q1199-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-240-242

Quant Time: Jan 29 00:39:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306518	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	270799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	149816	57.096	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.200%			
35) Dibromofluoromethane	8.165	113	111820	52.585	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.160%			
50) Toluene-d8	10.565	98	373062	49.377	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.760%			
62) 4-Bromofluorobenzene	12.847	95	117423	45.434	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.860%			
Target Compounds						
16) Acetone	4.441	43	3317	3.912	ug/l	95
44) Trichloroethene	9.353	130	27023	12.946	ug/l	89
64) Tetrachloroethene	11.100	164	1751	0.948	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085538.D
Acq On : 28 Jan 2025 15:01
Operator : JC\MD
Sample : Q1199-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-240-242

Quant Time: Jan 29 00:39:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

