

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085539.D
 Acq On : 28 Jan 2025 15:25
 Operator : JC\MD
 Sample : Q1199-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-192-GW-220-222

Quant Time: Jan 29 00:40:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	320348	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	257521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	96171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	154961	54.673	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.340%
35) Dibromofluoromethane	8.165	113	117804	53.007	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.020%
50) Toluene-d8	10.565	98	369236	46.761	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.520%
62) 4-Bromofluorobenzene	12.847	95	111786	41.385	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	82.780%
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	4790	5.230	ug/l	90
44) Trichloroethene	9.347	130	2299	1.054	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085539.D
Acq On : 28 Jan 2025 15:25
Operator : JC\MD
Sample : Q1199-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-192-GW-220-222

Quant Time: Jan 29 00:40:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

