

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085541.D
 Acq On : 28 Jan 2025 16:13
 Operator : JC\MD
 Sample : PB166233ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166233ZHE#11

Quant Time: Jan 29 00:40:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	172653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	321903	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	115243	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	154566	55.460	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 110.920%			
35) Dibromofluoromethane	8.165	113	116881	52.338	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.680%			
50) Toluene-d8	10.565	98	429295	54.104	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 108.200%			
62) 4-Bromofluorobenzene	12.847	95	128336	47.283	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 94.560%			
Target Compounds						
16) Acetone	4.424	43	59915	66.536	ug/l	96
43) Isopropyl Acetate	8.688	43	132541	26.146	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085541.D
Acq On : 28 Jan 2025 16:13
Operator : JC\MD
Sample : PB166233ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166233ZHE#11

Quant Time: Jan 29 00:40:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

