

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
 Data File : VN085543.D
 Acq On : 28 Jan 2025 17:01
 Operator : JC\MD
 Sample : PB166233ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166233ZHE#13

Quant Time: Jan 29 00:41:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	319233	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	299575	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	107137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150448	57.261	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.520%	
35) Dibromofluoromethane	8.165	113	111289	50.250	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.500%	
50) Toluene-d8	10.565	98	411098	52.244	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	104.480%	
62) 4-Bromofluorobenzene	12.847	95	121624	45.185	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	90.360%	
Target Compounds						
16) Acetone	4.430	43	58527	68.942	ug/l	94
43) Isopropyl Acetate	8.688	43	157501	31.330	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012825\
Data File : VN085543.D
Acq On : 28 Jan 2025 17:01
Operator : JC\MD
Sample : PB166233ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166233ZHE#13

Quant Time: Jan 29 00:41:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

