

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065672.D
 Acq On : 29 Jan 2021 12:48
 Operator : JC/MD
 Sample : VN0129MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0129MBS01

Quant Time: Jan 30 01:01:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	146461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	259325	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	231822	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	94586	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	90098	46.76	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.52%
35) Dibromofluoromethane	7.547	113	75048	48.84	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.68%
50) Toluene-d8	10.055	98	286865	47.49	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.98%
62) 4-Bromofluorobenzene	12.370	95	104586	48.08	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.16%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.836	85	29713	18.11	ug/l 99
3) Chloromethane	2.036	50	45652	18.94	ug/l 99
4) Vinyl Chloride	2.168	62	45128	18.49	ug/l 98
5) Bromomethane	2.537	94	27974	19.80	ug/l 96
6) Chloroethane	2.676	64	25841	18.30	ug/l 100
7) Trichlorofluoromethane	2.991	101	46625	18.27	ug/l 99
8) Diethyl Ether	3.377	74	23136	19.02	ug/l 92
9) 1,1,2-Trichlorotrifluo...	3.717	101	28924	18.67	ug/l 94
10) Methyl Iodide	3.904	142	41881	18.45	ug/l 98
11) Tert butyl alcohol	4.762	59	31339	86.83	ug/l 99
12) 1,1-Dichloroethene	3.695	96	30789	17.27	ug/l 93
13) Acrolein	3.573	56	22918	86.25	ug/l 94
14) Allyl chloride	4.274	41	61221	18.52	ug/l 90
15) Acrylonitrile	4.939	53	88283	93.89	ug/l 99
16) Acetone	3.782	43	63873	85.15	ug/l 96
17) Carbon Disulfide	4.004	76	88405	15.75	ug/l 99
18) Methyl Acetate	4.287	43	41600	19.79	ug/l 96
19) Methyl tert-butyl Ether	4.997	73	110499	18.57	ug/l 96
20) Methylene Chloride	4.499	84	37469	18.04	ug/l 93
21) trans-1,2-Dichloroethene	4.984	96	34039	17.96	ug/l 91
22) Diisopropyl ether	5.907	45	128466	19.01	ug/l 97
23) Vinyl Acetate	5.846	43	509205	94.35	ug/l 97
24) 1,1-Dichloroethane	5.795	63	67971	18.87	ug/l 97
25) 2-Butanone	6.794	43	111486	94.79	ug/l 94
26) 2,2-Dichloropropane	6.772	77	50140	17.40	ug/l 98
27) cis-1,2-Dichloroethene	6.782	96	40546	18.67	ug/l 93
28) Bromochloromethane	7.145	49	25469	16.83	ug/l 84
29) Tetrahydrofuran	7.174	42	78842	92.42	ug/l 94
30) Chloroform	7.328	83	62293	18.78	ug/l 98
31) Cyclohexane	7.608	56	62031	17.12	ug/l 95
32) 1,1,1-Trichloroethane	7.528	97	48365	17.91	ug/l 98
36) 1,1-Dichloropropene	7.749	75	48505	18.59	ug/l 97
37) Ethyl Acetate	6.884	43	48783	19.49	ug/l 98
38) Carbon Tetrachloride	7.730	117	40054	18.64	ug/l 100
39) Methylcyclohexane	9.042	83	53550	16.82	ug/l 92
40) Benzene	8.000	78	154416	19.38	ug/l 98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065672.D
 Acq On : 29 Jan 2021 12:48
 Operator : JC/MD
 Sample : VN0129MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0129MBS01

Quant Time: Jan 30 01:01:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.126	41	26356	19.61	ug/l #	100
42) 1,2-Dichloroethane	8.084	62	48796	20.31	ug/l	99
43) Isopropyl Acetate	8.126	43	84280	18.98	ug/l #	91
44) Trichloroethene	8.798	130	36533	18.89	ug/l	97
45) 1,2-Dichloropropane	9.081	63	43223	19.71	ug/l	99
46) Dibromomethane	9.171	93	24091	19.89	ug/l	93
47) Bromodichloromethane	9.367	83	48425	18.64	ug/l	99
48) Methyl methacrylate	9.164	41	39815	18.91	ug/l	91
49) 1,4-Dioxane	9.174	88	13091	384.15	ug/l	90
51) 4-Methyl-2-Pentanone	9.952	43	237793	97.17	ug/l	100
52) Toluene	10.122	92	91182	19.32	ug/l	98
53) t-1,3-Dichloropropene	10.351	75	56578	18.70	ug/l	97
54) cis-1,3-Dichloropropene	9.804	75	63840	18.94	ug/l	97
55) 1,1,2-Trichloroethane	10.531	97	36149	20.08	ug/l	94
56) Ethyl methacrylate	10.399	69	56224	18.88	ug/l	94
57) 1,3-Dichloropropane	10.675	76	63749	20.12	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.659	63	127805	92.94	ug/l	94
59) 2-Hexanone	10.720	43	163734	96.37	ug/l	100
60) Dibromochloromethane	10.871	129	34327	18.65	ug/l	99
61) 1,2-Dibromoethane	10.974	107	35677	19.88	ug/l	100
64) Tetrachloroethene	10.598	164	35559	18.30	ug/l	96
65) Chlorobenzene	11.402	112	92785	19.13	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.476	131	32520	18.97	ug/l	98
67) Ethyl Benzene	11.479	91	167275	18.36	ug/l	96
68) m/p-Xylenes	11.588	106	124689	37.80	ug/l	99
69) o-Xylene	11.916	106	60987	19.08	ug/l	98
70) Styrene	11.933	104	99758	18.96	ug/l	99
71) Bromoform	12.097	173	21134	17.83	ug/l #	98
73) Isopropylbenzene	12.219	105	154686	18.68	ug/l	99
74) N-amyl acetate	12.039	43	66491	18.98	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.473	83	47662	19.49	ug/l	98
76) 1,2,3-Trichloropropane	12.524	75	63600	25.71	ug/l #	100
77) Bromobenzene	12.495	156	34579	19.28	ug/l	74
78) n-propylbenzene	12.560	91	178156	18.25	ug/l	96
79) 2-Chlorotoluene	12.643	91	106894	18.44	ug/l	97
80) 1,3,5-Trimethylbenzene	12.701	105	131307	18.62	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.273	75	17170	17.78	ug/l	90
82) 4-Chlorotoluene	12.743	91	109878	18.57	ug/l	94
83) tert-Butylbenzene	12.965	119	107187	18.23	ug/l	99
84) 1,2,4-Trimethylbenzene	13.010	105	131211	18.66	ug/l	100
85) sec-Butylbenzene	13.141	105	140629	17.72	ug/l	98
86) p-Isopropyltoluene	13.257	119	125006	17.83	ug/l	99
87) 1,3-Dichlorobenzene	13.254	146	59516	18.62	ug/l	97
88) 1,4-Dichlorobenzene	13.334	146	60264	18.58	ug/l	96
89) n-Butylbenzene	13.585	91	108790	17.28	ug/l	98
90) Hexachloroethane	13.842	117	20773	16.94	ug/l	92
91) 1,2-Dichlorobenzene	13.624	146	61147	19.28	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.238	75	8406	17.32	ug/l	89
93) 1,2,4-Trichlorobenzene	14.878	180	28335	17.12	ug/l	97
94) Hexachlorobutadiene	14.977	225	13552	17.43	ug/l	99
95) Naphthalene	15.100	128	89502	17.05	ug/l	100
96) 1,2,3-Trichlorobenzene	15.280	180	29428	17.51	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
Data File : VN065672.D
Acq On : 29 Jan 2021 12:48
Operator : JC/MD
Sample : VN0129MBS01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01

Quant Time: Jan 30 01:01:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.624 min Scan# 1830
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:168 Resp: 146461
Ion Ratio Lower Upper
168 100
99 56.1 42.6 63.8



#2
Dichlorodifluoromethane
Concen: 18.11 ug/l
RT: 1.836 min Scan# 30
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 85 Resp: 29713
Ion Ratio Lower Upper
85 100
87 33.6 16.6 49.8









#3
Chloromethane
Concen: 18.94 ug/l
RT: 2.036 min Scan# 92
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 50 Resp: 45652
Ion Ratio Lower Upper
50 100
52 32.9 25.8 38.8



#4
Vinyl Chloride
Concen: 18.49 ug/l
RT: 2.168 min Scan# 133
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 62 Resp: 45128
Ion Ratio Lower Upper
62 100
64 32.3 25.1 37.7









#5
Bromomethane
Concen: 19.80 ug/l
RT: 2.537 min Scan# 248
Delta R.T. -0.004 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 94 Resp: 27974
Ion Ratio Lower Upper
94 100
96 94.6 73.0 109.4



#6
Chloroethane
Concen: 18.30 ug/l
RT: 2.676 min Scan# 291
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 64 Resp: 25841
Ion Ratio Lower Upper
64 100
66 31.9 25.4 38.2









#7
Trichlorofluoromethane
Concen: 18.27 ug/l
RT: 2.991 min Scan# 389
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:101 Resp: 46625
Ion Ratio Lower Upper
101 100
103 64.2 52.1 78.1



#8
Diethyl Ether
Concen: 19.02 ug/l
RT: 3.377 min Scan# 509
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 74 Resp: 23136
Ion Ratio Lower Upper
74 100
45 100.1 46.4 139.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.67 ug/l
RT: 3.717 min Scan# 615
Delta R.T. -0.004 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:101 Resp: 28924
Ion Ratio Lower Upper
101 100
85 42.5 34.6 51.8
151 71.8 63.5 95.3



#10
Methyl Iodide
Concen: 18.45 ug/l
RT: 3.904 min Scan# 673
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:142 Resp: 41881
Ion Ratio Lower Upper
142 100
127 38.1 31.7 47.5
141 13.9 10.9 16.3









#11
Tert butyl alcohol
Concen: 86.83 ug/l
RT: 4.762 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 59 Resp: 31339
Ion Ratio Lower Upper
59 100
57 9.6 8.0 12.0



#12
1,1-Dichloroethene
Concen: 17.27 ug/l
RT: 3.695 min Scan# 608
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 96 Resp: 30789
Ion Ratio Lower Upper
96 100
61 163.5 122.5 183.7
98 61.2 50.8 76.2









#13
Acrolein
Concen: 86.25 ug/l
RT: 3.573 min Scan# 570
Delta R.T. 0.004 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 56 Resp: 22918
Ion Ratio Lower Upper
56 100
55 74.9 56.0 84.0



#14
Allyl chloride
Concen: 18.52 ug/l
RT: 4.274 min Scan# 788
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 41 Resp: 61221
Ion Ratio Lower Upper
41 100
39 56.8 53.0 79.4
76 32.2 28.9 43.3









#15
Acrylonitrile
Concen: 93.89 ug/l
RT: 4.939 min Scan# 995
Delta R.T. 0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 53	Resp: 88283
Ion Ratio	Lower Upper
53 100	
52 82.3	66.1 99.1
51 34.3	27.9 41.9



#16
Acetone
Concen: 85.15 ug/l
RT: 3.782 min Scan# 635
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 43	Resp: 63873
Ion Ratio	Lower Upper
43 100	
58 35.9	27.0 40.6









#17
Carbon Disulfide
Concen: 15.75 ug/l
RT: 4.004 min Scan# 704
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 76 Resp: 88405
Ion Ratio Lower Upper
76 100
78 9.4 7.8 11.6



#18
Methyl Acetate
Concen: 19.79 ug/l
RT: 4.287 min Scan# 792
Delta R.T. 0.004 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 43 Resp: 41600
Ion Ratio Lower Upper
43 100
74 22.3 19.5 29.3









#19
Methyl tert-butyl Ether
Concen: 18.57 ug/l
RT: 4.997 min Scan# 1013
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 73 Resp: 110499
Ion Ratio Lower Upper
73 100
57 23.5 17.3 25.9



#20
Methylene Chloride
Concen: 18.04 ug/l
RT: 4.499 min Scan# 858
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 84 Resp: 37469
Ion Ratio Lower Upper
84 100
49 129.4 94.6 141.8
51 41.1 29.9 44.9
86 63.7 51.4 77.2









#21
trans-1,2-Dichloroethene
Concen: 17.96 ug/l
RT: 4.984 min Scan# 1009
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 96 Resp: 34039
Ion Ratio Lower Upper
96 100
61 144.4 104.9 157.3
98 60.6 51.8 77.8



#22
Diisopropyl ether
Concen: 19.01 ug/l
RT: 5.907 min Scan# 1296
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 45 Resp: 128466
Ion Ratio Lower Upper
45 100
43 59.9 48.9 73.3
87 25.9 23.7 35.5
59 12.2 9.9 14.9









#23
 Vinyl Acetate
 Concen: 94.35 ug/l
 RT: 5.846 min Scan# 1277
 Delta R.T. -0.000 min
 Lab File: VN065672.D
 Acq: 29 Jan 2021 12:48




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0129MBS01


Tgt Ion: 43 Resp: 509205
 Ion Ratio Lower Upper
 43 100
 86 9.7 8.6 13.0



#24
 1,1-Dichloroethane
 Concen: 18.87 ug/l
 RT: 5.795 min Scan# 1261
 Delta R.T. -0.003 min
 Lab File: VN065672.D
 Acq: 29 Jan 2021 12:48

Tgt Ion: 63 Resp: 67971
 Ion Ratio Lower Upper
 63 100
 98 5.7 3.5 10.4
 100 3.6 2.3 6.9









#25
2-Butanone
Concen: 94.79 ug/l
RT: 6.794 min Scan# 1572
Delta R.T. 0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampled :
VN0129MBS01


Tgt Ion: 43 Resp: 111486
Ion Ratio Lower Upper
43 100
72 25.5 23.1 34.7



#26
2,2-Dichloropropane
Concen: 17.40 ug/l
RT: 6.772 min Scan# 1565
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 77 Resp: 50140
Ion Ratio Lower Upper
77 100
97 24.7 11.9 35.9









#27
cis-1,2-Dichloroethene
Concen: 18.67 ug/l
RT: 6.782 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampled :
VN0129MBS01


Tgt Ion:	96	Resp:	40546
Ion	Ratio	Lower	Upper
96	100		
61	149.1	0.0	276.8
98	63.3	0.0	129.4



#28
Bromochloromethane
Concen: 16.83 ug/l
RT: 7.145 min Scan# 1681
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:	49	Resp:	25469
Ion	Ratio	Lower	Upper
49	100		
129	0.2	0.0	3.6
130	70.4	67.8	101.6





#29

Tetrahydrofuran

Concen: 92.42 ug/l

RT: 7.174 min Scan# 1690





Delta R.T. 0.003 min

Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion: 42 Resp: 78842

Ion	Ratio	Lower	Upper
42	100		
72	41.5	36.6	55.0
71	39.3	34.3	51.5



#30

Chloroform

Concen: 18.78 ug/l

RT: 7.328 min Scan# 1738




Delta R.T. 0.000 min


Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion: 83 Resp: 62293

Ion	Ratio	Lower	Upper
83	100		
85	64.2	52.6	78.8





#31

Cyclohexane

Concen: 17.12 ug/l

RT: 7.608 min Scan# 1825





Delta R.T. -0.003 min

Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion: 56 Resp: 62031

Ion	Ratio	Lower	Upper
56	100		
69	36.1	26.6	39.8
84	84.0	70.4	105.6



#32

1,1,1-Trichloroethane

Concen: 17.91 ug/l

RT: 7.528 min Scan# 1800




Delta R.T. 0.004 min


Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion: 97 Resp: 48365

Ion	Ratio	Lower	Upper
97	100		
99	64.3	50.8	76.2
61	45.4	34.3	51.5









#33
1,2-Dichloroethane-d4
Concen: 46.76 ug/l
RT: 7.984 min Scan# 1942
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 65 Resp: 90098
Ion Ratio Lower Upper
65 100
67 55.9 0.0 107.2



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.547 min Scan# 2117
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 114 Resp: 259325
Ion Ratio Lower Upper
114 100
63 22.2 0.0 38.6
88 16.4 0.0 30.6





#35

Dibromofluoromethane

Concen: 48.84 ug/l

RT: 7.547 min Scan# 1806





Delta R.T. -0.000 min

Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion:	113	Resp:	75048
Ion Ratio	Lower	Upper	
113	100		
111	103.6	82.2	123.2
192	17.5	15.4	23.0

Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01



#36

1,1-Dichloropropene

Concen: 18.59 ug/l




RT: 7.749 min Scan# 1869


Delta R.T. 0.000 min

Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion:	75	Resp:	48505
Ion Ratio	Lower	Upper	
75	100		
110	33.4	18.2	54.6
77	31.2	24.9	37.3





#37

Ethyl Acetate

Concen: 19.49 ug/l

RT: 6.884 min Scan# 1600





Delta R.T. 0.000 min

Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01

Tgt Ion: 43	Resp: 48783
Ion Ratio	Lower Upper
43 100	
61 15.4	11.6 17.4
70 11.0	8.6 13.0



#38

Carbon Tetrachloride

Concen: 18.64 ug/l




RT: 7.730 min Scan# 1863


Delta R.T. 0.000 min

Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion:117	Resp:	40054	
Ion	Ratio	Lower	Upper
117	100		
119	94.9	76.0	114.0
121	31.4	24.5	36.7









#39
Methylcyclohexane
Concen: 16.82 ug/l
RT: 9.042 min Scan# 2271
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 83	Resp: 53550
Ion Ratio	Lower Upper
83 100	
55 82.9	58.9 88.3
98 43.7	35.9 53.9



#40
Benzene
Concen: 19.38 ug/l
RT: 8.000 min Scan# 1947
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 78	Resp: 154416
Ion Ratio	Lower Upper
78 100	
77 22.9	19.0 28.4









#41
Methacrylonitrile
Concen: 19.61 ug/l
RT: 7.126 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:	41	Resp:	26356
Ion	Ratio	Lower	Upper
41	100		
39	51.2	0.0	0.0#
67	75.8	0.0	0.0#
52	27.5	0.0	0.0#



#42
1,2-Dichloroethane
Concen: 20.31 ug/l
RT: 8.084 min Scan# 1973
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:	62	Resp:	48796
Ion	Ratio	Lower	Upper
62	100		
98	9.6	0.0	18.6









#43
Isopropyl Acetate
Concen: 18.98 ug/l
RT: 8.126 min Scan# 1986
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 43 Resp: 84280
Ion Ratio Lower Upper
43 100
61 22.2 22.5 33.7#
87 12.5 11.1 16.7



#44
Trichloroethene
Concen: 18.89 ug/l
RT: 8.798 min Scan# 2195
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:130 Resp: 36533
Ion Ratio Lower Upper
130 100
95 102.8 0.0 199.6









#45
1,2-Dichloropropane
Concen: 19.71 ug/l
RT: 9.081 min Scan# 2283
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 63 Resp: 43223
Ion Ratio Lower Upper
63 100
65 31.9 24.9 37.3



#46
Dibromomethane
Concen: 19.89 ug/l
RT: 9.171 min Scan# 2311
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 93 Resp: 24091
Ion Ratio Lower Upper
93 100
95 83.3 67.0 100.6
174 90.9 83.2 124.8









#47
Bromodichloromethane
Concen: 18.64 ug/l
RT: 9.367 min Scan# 2372
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 83 Resp: 48425
Ion Ratio Lower Upper
83 100
85 64.7 51.0 76.4
127 7.8 7.0 10.6



#48
Methyl methacrylate
Concen: 18.91 ug/l
RT: 9.164 min Scan# 2309
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 41 Resp: 39815
Ion Ratio Lower Upper
41 100
69 88.4 75.9 113.9
39 48.4 46.6 69.8









#49
1,4-Dioxane
Concen: 384.15 ug/l
RT: 9.174 min Scan# 2312
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 88 Resp: 13091
Ion Ratio Lower Upper
88 100
43 31.3 24.7 37.1
58 76.8 52.6 78.8



#50
Toluene-d8
Concen: 47.49 ug/l
RT: 10.055 min Scan# 2586
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 98 Resp: 286865
Ion Ratio Lower Upper
98 100
100 64.7 50.5 75.7









#51
4-Methyl-2-Pentanone
Concen: 97.17 ug/l
RT: 9.952 min Scan# 2554
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 43 Resp: 237793
Ion Ratio Lower Upper
43 100
58 40.8 32.5 48.7



#52
Toluene
Concen: 19.32 ug/l
RT: 10.122 min Scan# 2607
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 92 Resp: 91182
Ion Ratio Lower Upper
92 100
91 172.5 140.7 211.1









#53
t-1,3-Dichloropropene
Concen: 18.70 ug/l
RT: 10.351 min Scan# 2678
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 75 Resp: 56578
Ion Ratio Lower Upper
75 100
77 30.9 26.2 39.4



#54
cis-1,3-Dichloropropene
Concen: 18.94 ug/l
RT: 9.804 min Scan# 2508
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 75 Resp: 63840
Ion Ratio Lower Upper
75 100
77 31.8 24.9 37.3
39 42.6 36.2 54.4









#55
1,1,2-Trichloroethane
Concen: 20.08 ug/l
RT: 10.531 min Scan# 2734
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:	97	Resp:	36149
Ion	Ratio	Lower	Upper
97	100		
83	91.3	67.7	101.5
85	60.0	44.5	66.7
99	64.6	49.4	74.0



#56
Ethyl methacrylate
Concen: 18.88 ug/l
RT: 10.399 min Scan# 2693
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:	69	Resp:	56224
Ion	Ratio	Lower	Upper
69	100		
41	65.2	49.6	74.4
39	30.1	28.0	42.0









#57
1,3-Dichloropropane
Concen: 20.12 ug/l
RT: 10.675 min Scan# 2779
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 76 Resp: 63749
Ion Ratio Lower Upper
76 100
78 32.0 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 92.94 ug/l
RT: 9.659 min Scan# 2463
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 63 Resp: 127805
Ion Ratio Lower Upper
63 100
106 24.8 22.6 33.8









#59
2-Hexanone
Concen: 96.37 ug/l
RT: 10.720 min Scan# 2793
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampled :
VN0129MBS01


Tgt Ion: 43 Resp: 163734
Ion Ratio Lower Upper
43 100
58 56.5 28.1 84.3



#60
Dibromochloromethane
Concen: 18.65 ug/l
RT: 10.871 min Scan# 2840
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 129 Resp: 34327
Ion Ratio Lower Upper
129 100
127 77.6 38.5 115.5









#61
1,2-Dibromoethane
Concen: 19.88 ug/l
RT: 10.974 min Scan# 2872
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:107 Resp: 35677
Ion Ratio Lower Upper
107 100
109 94.7 75.6 113.4



#62
4-Bromofluorobenzene
Concen: 48.08 ug/l
RT: 12.370 min Scan# 3306
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 95 Resp: 104586
Ion Ratio Lower Upper
95 100
174 69.9 0.0 169.0
176 65.9 0.0 162.4









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.376 min Scan# 2997
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:117 Resp: 231822
Ion Ratio Lower Upper
117 100
82 57.8 43.7 65.5
119 32.4 26.1 39.1



#64
Tetrachloroethene
Concen: 18.30 ug/l
RT: 10.598 min Scan# 2755
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:164 Resp: 35559
Ion Ratio Lower Upper
164 100
166 129.6 102.6 154.0
129 101.3 76.2 114.4
131 97.6 73.4 110.2









#65
Chlorobenzene
Concen: 19.13 ug/l
RT: 11.402 min Scan# 3005
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:112 Resp: 92785
Ion Ratio Lower Upper
112 100
114 32.1 26.1 39.1



#66
1,1,1,2-Tetrachloroethane
Concen: 18.97 ug/l
RT: 11.476 min Scan# 3028
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:131 Resp: 32520
Ion Ratio Lower Upper
131 100
133 95.2 48.0 144.0
119 71.2 34.1 102.2









#67
Ethyl Benzene
Concen: 18.36 ug/l
RT: 11.479 min Scan# 3029
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 91 Resp: 167275
Ion Ratio Lower Upper
91 100
106 29.8 25.5 38.3



#68
m/p-Xylenes
Concen: 37.80 ug/l
RT: 11.588 min Scan# 3063
Delta R.T. -0.004 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 106 Resp: 124689
Ion Ratio Lower Upper
106 100
91 203.9 161.4 242.0









#69
o-Xylene
Concen: 19.08 ug/l
RT: 11.916 min Scan# 3165
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:106 Resp: 60987
Ion Ratio Lower Upper
106 100
91 217.4 106.8 320.4



#70
Styrene
Concen: 18.96 ug/l
RT: 11.933 min Scan# 3170
Delta R.T. 0.001 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:104 Resp: 99758
Ion Ratio Lower Upper
104 100
78 49.7 39.0 58.4
103 54.6 44.5 66.7









#71
Bromoform
Concen: 17.83 ug/l
RT: 12.097 min Scan# 3221
Delta R.T. 0.001 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:173 Resp: 21134
Ion Ratio Lower Upper
173 100
175 48.9 23.8 71.2
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.312 min Scan# 3599
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:152 Resp: 94586
Ion Ratio Lower Upper
152 100
115 60.8 28.5 85.6
150 166.0 0.0 347.6









#73
Isopropylbenzene
Concen: 18.68 ug/l
RT: 12.219 min Scan# 3259
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:105 Resp: 154686
Ion Ratio Lower Upper
105 100
120 25.9 13.3 39.9



#74
N-ethyl acetate
Concen: 18.98 ug/l
RT: 12.039 min Scan# 3203
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 43 Resp: 66491
Ion Ratio Lower Upper
43 100
70 41.8 34.4 51.6
55 24.0 20.7 31.1
61 24.8 19.1 28.7









#75
1,1,2,2-Tetrachloroethane
Concen: 19.49 ug/l
RT: 12.473 min Scan# 3338
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 83 Resp: 47662
Ion Ratio Lower Upper
83 100
131 9.0 5.3 15.8
85 62.8 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 25.71 ug/l
RT: 12.524 min Scan# 3354
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 75 Resp: 63600
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 19.28 ug/l

RT: 12.495 min Scan# 3345





Delta R.T. 0.000 min

Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion:	156	Resp:	34579
Ion	Ratio	Lower	Upper
156	100		
77	248.9	96.3	288.8
158	100.6	49.0	147.0

Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01



#78

n-propylbenzene

Concen: 18.25 ug/l




RT: 12.560 min Scan# 3365


Delta R.T. -0.004 min

Lab File: VN065672.D

Acq: 29 Jan 2021 12:48

Tgt Ion:	91	Resp:	178156
Ion	Ratio	Lower	Upper
91	100		
120	21.5	11.8	35.3









#79
2-Chlorotoluene
Concen: 18.44 ug/l
RT: 12.643 min Scan# 3391
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 91 Resp: 106894
Ion Ratio Lower Upper
91 100
126 33.1 17.4 52.3



#80
1,3,5-Trimethylbenzene
Concen: 18.62 ug/l
RT: 12.701 min Scan# 3409
Delta R.T. -0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 105 Resp: 131307
Ion Ratio Lower Upper
105 100
120 48.9 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 17.78 ug/l
RT: 12.273 min Scan# 3276
Delta R.T. 0.003 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 75 Resp: 17170
Ion Ratio Lower Upper
75 100
53 82.7 74.6 111.8
89 38.7 35.8 53.8



#82
4-Chlorotoluene
Concen: 18.57 ug/l
RT: 12.743 min Scan# 3422
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 91 Resp: 109878
Ion Ratio Lower Upper
91 100
126 31.2 17.2 51.6









#83
tert-Butylbenzene
Concen: 18.23 ug/l
RT: 12.965 min Scan# 3491
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:119 Resp: 107187
Ion Ratio Lower Upper
119 100
91 65.8 33.0 98.9
134 24.3 12.7 38.1



#84
1,2,4-Trimethylbenzene
Concen: 18.66 ug/l
RT: 13.010 min Scan# 3505
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:105 Resp: 131211
Ion Ratio Lower Upper
105 100
120 44.8 22.4 67.0









#85
sec-Butylbenzene
Concen: 17.72 ug/l
RT: 13.141 min Scan# 3546
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:105 Resp: 140629
Ion Ratio Lower Upper
105 100
134 19.2 10.0 30.0



#86
p-Isopropyltoluene
Concen: 17.83 ug/l
RT: 13.257 min Scan# 3582
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:119 Resp: 125006
Ion Ratio Lower Upper
119 100
134 26.0 13.3 39.8
91 23.4 11.8 35.3









#87

1,3-Dichlorobenzene
Concen: 18.62 ug/l
RT: 13.254 min Scan# 3581
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01




Tgt Ion:146 Resp: 59516
Ion Ratio Lower Upper
146 100
111 43.4 19.5 58.5
148 64.7 32.3 96.9




#88

1,4-Dichlorobenzene
Concen: 18.58 ug/l
RT: 13.334 min Scan# 3606
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:146 Resp: 60264
Ion Ratio Lower Upper
146 100
111 43.2 19.0 57.0
148 64.0 31.8 95.3









#89
n-Butylbenzene
Concen: 17.28 ug/l
RT: 13.585 min Scan# 3684
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion: 91 Resp: 108790
Ion Ratio Lower Upper
91 100
92 52.4 26.1 78.3
134 23.4 13.0 38.9



#90
Hexachloroethane
Concen: 16.94 ug/l
RT: 13.842 min Scan# 3764
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 117 Resp: 20773
Ion Ratio Lower Upper
117 100
201 68.9 37.8 113.3









#91
1,2-Dichlorobenzene
Concen: 19.28 ug/l
RT: 13.624 min Scan# 3696
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:146 Resp: 61147
Ion Ratio Lower Upper
146 100
111 43.9 20.1 60.3
148 63.5 32.0 96.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 17.32 ug/l
RT: 14.238 min Scan# 3887
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion: 75 Resp: 8406
Ion Ratio Lower Upper
75 100
155 78.2 44.1 132.4
157 99.9 56.0 168.2









#93
1,2,4-Trichlorobenzene
Concen: 17.12 ug/l
RT: 14.878 min Scan# 4086
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48




Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01


Tgt Ion:180 Resp: 28335
Ion Ratio Lower Upper
180 100
182 98.2 47.3 142.0
145 31.7 15.6 46.8



#94
Hexachlorobutadiene
Concen: 17.43 ug/l
RT: 14.977 min Scan# 4117
Delta R.T. 0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:225 Resp: 13552
Ion Ratio Lower Upper
225 100
223 64.5 31.9 95.9
227 63.6 31.6 95.0









#95
Naphthalene
Concen: 17.05 ug/l
RT: 15.100 min Scan# 4155
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Instrument :
MSVOA_N
ClientSampleId :
VN0129MBS01

Tgt Ion:128 Resp: 89502
Ion Ratio Lower Upper
128 100
127 12.9 10.2 15.2
129 10.7 8.6 13.0



#96
1,2,3-Trichlorobenzene
Concen: 17.51 ug/l
RT: 15.280 min Scan# 4211
Delta R.T. -0.000 min
Lab File: VN065672.D
Acq: 29 Jan 2021 12:48

Tgt Ion:180 Resp: 29428
Ion Ratio Lower Upper
180 100
182 98.0 47.4 142.3
145 33.0 16.5 49.5

