

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065680.D
 Acq On : 29 Jan 2021 16:28
 Operator : JC/MD
 Sample : M1288-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-2D(ONSITE)

Quant Time: Jan 30 01:03:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	203277	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	348558	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	304287	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	108283	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	124303	46.48	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.96%
35) Dibromofluoromethane	7.550	113	98940	47.90	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.80%
50) Toluene-d8	10.058	98	434732	53.54	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.08%
62) 4-Bromofluorobenzene	12.373	95	139465	47.70	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.40%
Target Compounds						
					Qvalue	
30) Chloroform	7.332	83	1096	0.24	ug/l	85
38) Carbon Tetrachloride	7.730	117	1897	0.66	ug/l	98
44) Trichloroethene	8.801	130	3016	1.16	ug/l	89
64) Tetrachloroethene	10.598	164	367596	144.15	ug/l	96
90) Hexachloroethane	13.842	117	11378	8.10	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
Data File : VN065680.D
Acq On : 29 Jan 2021 16:28
Operator : JC/MD
Sample : M1288-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-2D(ONSITE)

Quant Time: Jan 30 01:03:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

