

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065681.D
 Acq On : 29 Jan 2021 16:53
 Operator : JC/MD
 Sample : M1288-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-MW-5S

Quant Time: Jan 30 01:04:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	194109	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	332957	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	286811	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	103300	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	118701	46.48	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.96%
35) Dibromofluoromethane	7.550	113	96979	49.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.30%
50) Toluene-d8	10.058	98	412149	53.14	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.28%
62) 4-Bromofluorobenzene	12.373	95	133876	47.93	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.86%
Target Compounds						
27) cis-1,2-Dichloroethene	6.782	96	1928	0.67	ug/l	88
64) Tetrachloroethene	10.601	164	1551	0.65	ug/l	91
85) sec-Butylbenzene	13.141	105	4709	0.54	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
Data File : VN065681.D
Acq On : 29 Jan 2021 16:53
Operator : JC/MD
Sample : M1288-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-MW-5S

Quant Time: Jan 30 01:04:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

