

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065686.D
 Acq On : 29 Jan 2021 18:55
 Operator : JC/MD
 Sample : M1254-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CRT-15

Quant Time: Jan 30 01:06:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	192908	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	327475	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	287297	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	106685	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	117234	46.19	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.38%
35) Dibromofluoromethane	7.550	113	94470	48.68	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.36%
50) Toluene-d8	10.058	98	410278	53.78	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.56%
62) 4-Bromofluorobenzene	12.370	95	134989	49.14	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.28%
Target Compounds						
					Qvalue	
16) Acetone	3.782	43	31509	31.89	ug/l	99
20) Methylene Chloride	4.502	84	10986	4.02	ug/l	97
43) Isopropyl Acetate	8.132	43	16689	2.98	ug/l #	90
95) Naphthalene	15.100	128	4669	4.08	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
Data File : VN065686.D
Acq On : 29 Jan 2021 18:55
Operator : JC/MD
Sample : M1254-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CRT-15

Quant Time: Jan 30 01:06:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

