

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065687.D
 Acq On : 29 Jan 2021 19:19
 Operator : JC/MD
 Sample : M1254-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CRT-5

Quant Time: Jan 30 01:06:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	199850	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	345521	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	306732	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	119393	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	122709	46.67	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.34%
35) Dibromofluoromethane	7.547	113	98914	48.31	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.62%
50) Toluene-d8	10.058	98	431183	53.57	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.14%
62) 4-Bromofluorobenzene	12.373	95	144956	50.01	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.02%
Target Compounds						
16) Acetone	3.782	43	25949	25.35	ug/l	98
20) Methylene Chloride	4.499	84	10513	3.71	ug/l	97
43) Isopropyl Acetate	8.129	43	17280	2.92	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
Data File : VN065687.D
Acq On : 29 Jan 2021 19:19
Operator : JC/MD
Sample : M1254-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CRT-5

Quant Time: Jan 30 01:06:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

