

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065688.D
 Acq On : 29 Jan 2021 19:44
 Operator : JC/MD
 Sample : M1255-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CRT-3

Quant Time: Jan 30 01:06:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	189651	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	329095	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	288591	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	108294	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	118357	47.44	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.88%
35) Dibromofluoromethane	7.547	113	95353	48.89	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.78%
50) Toluene-d8	10.058	98	409157	53.37	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.74%
62) 4-Bromofluorobenzene	12.373	95	136046	49.28	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.56%
Target Compounds						
16) Acetone	3.782	43	25226	25.97	ug/l	99
20) Methylene Chloride	4.505	84	10882	4.05	ug/l	96
43) Isopropyl Acetate	8.129	43	16491	2.93	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
Data File : VN065688.D
Acq On : 29 Jan 2021 19:44
Operator : JC/MD
Sample : M1255-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CRT-3

Quant Time: Jan 30 01:06:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

