

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065690.D
 Acq On : 29 Jan 2021 20:33
 Operator : JC/MD
 Sample : M1255-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CRT-17

Quant Time: Jan 30 01:07:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:47:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	195322	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	338153	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	294131	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.312	152	112662	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	119858	46.64	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.28%
35) Dibromofluoromethane	7.547	113	97926	48.87	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.74%
50) Toluene-d8	10.058	98	416628	52.89	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.78%
62) 4-Bromofluorobenzene	12.370	95	139127	49.05	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.10%
Target Compounds						
16) Acetone	3.779	43	23860	23.85	ug/l	95
20) Methylene Chloride	4.499	84	10444	3.77	ug/l	97
43) Isopropyl Acetate	8.129	43	16737	2.89	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
Data File : VN065690.D
Acq On : 29 Jan 2021 20:33
Operator : JC/MD
Sample : M1255-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CRT-17

Quant Time: Jan 30 01:07:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:47:28 2021
Response via : Initial Calibration

