


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012921\
 Data File : VN065667.D
 Acq On : 29 Jan 2021 9:22
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012521W.M
 Title : SW846 8260
 Last Update : Mon Jan 25 13:47:28 2021



AutoFind: Scans 3305, 3306, 3307; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	4699	PASS
75	95	30	60	45.9	12852	PASS
95	95	100	100	100.0	28003	PASS
96	95	5	9	6.6	1856	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	66.5	18619	PASS
175	174	5	9	7.6	1406	PASS
176	174	95	101	95.8	17829	PASS
177	176	5	9	7.5	1330	PASS