

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012924\
 Data File : VN080789.D
 Acq On : 29 Jan 2024 14:12
 Operator : JC\MD
 Sample : VN0129WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0129WBL01

Quant Time: Jan 29 15:36:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	103508	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	246556	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	250302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	96236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	84038	46.931	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.860%
35) Dibromofluoromethane	8.165	113	67696	44.878	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.760%
50) Toluene-d8	10.571	98	272977	46.237	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.480%
62) 4-Bromofluorobenzene	12.847	95	108782	52.875	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012924\
Data File : VN080789.D
Acq On : 29 Jan 2024 14:12
Operator : JC\MD
Sample : VN0129WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0129WBL01

Quant Time: Jan 29 15:36:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

