

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012924\
 Data File : VN080803.D
 Acq On : 29 Jan 2024 20:05
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 30 01:32:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.743	0.722	2.8	74	0.00
3 P	Chloromethane	1.116	1.080	3.2	82	0.00
4 C	Vinyl Chloride	0.849	0.840	1.1#	84	0.00
5 T	Bromomethane	0.339	0.316	6.8	78	0.00
6 T	Chloroethane	0.452	0.483	-6.9	89	0.00
7 T	Trichlorofluoromethane	1.098	1.073	2.3	78	0.00
8 T	Diethyl Ether	0.507	0.512	-1.0	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.641	0.619	3.4	76	0.00
10 T	Methyl Iodide	0.541	0.516	4.6	72	0.00
11 T	Tert butyl alcohol	0.141	0.149	-5.7	89	0.00
12 CM	1,1-Dichloroethene	0.702	0.663	5.6#	79	0.00
13 T	Acrolein	0.178	0.194	-9.0	88	0.00
14 T	Allyl chloride	1.387	1.350	2.7	78	0.00
15 T	Acrylonitrile	0.424	0.453	-6.8	85	0.00
16 T	Acetone	0.314	0.313	0.3	82	0.00
17 T	Carbon Disulfide	2.185	2.072	5.2	76	0.00
18 T	Methyl Acetate	1.093	1.147	-4.9	83	0.00
19 T	Methyl tert-butyl Ether	2.473	2.479	-0.2	82	0.00
20 T	Methylene Chloride	0.823	0.795	3.4	80	0.00
21 T	trans-1,2-Dichloroethene	0.762	0.731	4.1	79	0.00
22 T	Diisopropyl ether	3.061	3.161	-3.3	83	0.00
23 T	Vinyl Acetate	2.215	2.455	-10.8	84	0.00
24 P	1,1-Dichloroethane	1.552	1.617	-4.2	82	0.00
25 T	2-Butanone	0.596	0.626	-5.0	87	0.00
26 T	2,2-Dichloropropane	1.255	1.195	4.8	78	0.00
27 T	cis-1,2-Dichloroethene	0.883	0.877	0.7	83	0.00
28 T	Bromochloromethane	0.822	0.855	-4.0	82	0.00
29 T	Tetrahydrofuran	0.403	0.433	-7.4	88	0.00
30 C	Chloroform	1.442	1.448	-0.4#	83	0.00
31 T	Cyclohexane	1.483	1.370	7.6	78	0.00
32 T	1,1,1-Trichloroethane	1.169	1.190	-1.8	82	0.00
33 S	1,2-Dichloroethane-d4	0.865	0.854	1.3	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.306	0.287	6.2	77	0.00
36 T	1,1-Dichloropropene	0.569	0.543	4.6	80	0.00
37 T	Ethyl Acetate	0.585	0.580	0.9	85	0.00
38 T	Carbon Tetrachloride	0.458	0.448	2.2	78	0.00
39 T	Methylcyclohexane	0.565	0.539	4.6	76	0.00
40 TM	Benzene	1.671	1.674	-0.2	83	0.00
41 T	Methacrylonitrile	0.345	0.358	-3.8	93	0.00
42 TM	1,2-Dichloroethane	0.544	0.563	-3.5	83	0.00
43 T	Isopropyl Acetate	1.066	1.082	-1.5	87	0.00
44 TM	Trichloroethene	0.368	0.349	5.2	82	0.00
45 C	1,2-Dichloropropane	0.452	0.467	-3.3#	83	0.00
46 T	Dibromomethane	0.268	0.269	-0.4	85	0.00
47 T	Bromodichloromethane	0.544	0.529	2.8	79	0.00
48 T	Methyl methacrylate	0.517	0.493	4.6	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012924\
 Data File : VN080803.D
 Acq On : 29 Jan 2024 20:05
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 30 01:32:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 29 12:54:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	87	0.00
50 S	Toluene-d8	1.197	1.148	4.1	77	0.00
51 T	4-Methyl-2-Pentanone	0.596	0.626	-5.0	86	0.00
52 CM	Toluene	0.958	0.974	-1.7#	83	0.00
53 T	t-1,3-Dichloropropene	0.634	0.636	-0.3	80	0.00
54 T	cis-1,3-Dichloropropene	0.690	0.670	2.9	79	0.00
55 T	1,1,2-Trichloroethane	0.368	0.370	-0.5	84	0.00
56 T	Ethyl methacrylate	0.566	0.637	-12.5	82	0.00
57 T	1,3-Dichloropropane	0.693	0.710	-2.5	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.330	0.345	-4.5	85	0.00
59 T	2-Hexanone	0.463	0.463	0.0	82	0.00
60 T	Dibromochloromethane	0.343	0.347	-1.2	81	0.00
61 T	1,2-Dibromoethane	0.345	0.353	-2.3	81	0.00
62 S	4-Bromofluorobenzene	0.417	0.395	5.3	78	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.329	0.325	1.2	83	0.00
65 PM	Chlorobenzene	1.091	1.052	3.6	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.355	0.363	-2.3	81	0.00
67 C	Ethyl Benzene	2.028	2.066	-1.9#	83	0.00
68 T	m/p-Xylenes	0.720	0.747	-3.8	82	0.00
69 T	o-Xylene	0.716	0.752	-5.0	86	0.00
70 T	Styrene	1.149	1.231	-7.1	83	0.00
71 P	Bromoform	0.240	0.246	-2.5	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	4.579	4.448	2.9	81	0.00
74 T	N-amyl acetate	2.843	2.813	1.1	84	0.00
75 P	1,1,2,2-Tetrachloroethane	1.611	1.513	6.1	84	0.00
76 T	1,2,3-Trichloropropane	1.508	1.420	5.8	74	0.00
77 T	Bromobenzene	0.979	0.907	7.4	80	0.00
78 T	n-propylbenzene	5.470	5.560	-1.6	82	0.00
79 T	2-Chlorotoluene	3.477	3.323	4.4	82	0.00
80 T	1,3,5-Trimethylbenzene	3.647	3.640	0.2	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.585	0.554	5.3	77	0.00
82 T	4-Chlorotoluene	3.443	3.318	3.6	83	0.00
83 T	tert-Butylbenzene	2.912	2.859	1.8	81	0.00
84 T	1,2,4-Trimethylbenzene	3.695	3.611	2.3	81	0.00
85 T	sec-Butylbenzene	4.293	4.313	-0.5	81	0.00
86 T	p-Isopropyltoluene	3.336	3.266	2.1	78	0.00
87 T	1,3-Dichlorobenzene	1.751	1.679	4.1	83	0.00
88 T	1,4-Dichlorobenzene	1.755	1.642	6.4	80	0.00
89 T	n-Butylbenzene	3.265	3.127	4.2	76	0.00
90 T	Hexachloroethane	0.584	0.558	4.5	80	0.00
91 T	1,2-Dichlorobenzene	1.602	1.584	1.1	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.301	0.297	1.3	83	0.00
93 T	1,2,4-Trichlorobenzene	0.825	0.751	9.0	78	0.00
94 T	Hexachlorobutadiene	0.360	0.306	15.0	78	0.00
95 T	Naphthalene	3.156	2.954	6.4	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012924\
Data File : VN080803.D
Acq On : 29 Jan 2024 20:05
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 30 01:32:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N012924W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 29 12:54:23 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.811	0.730	10.0	76	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6