

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
 Data File : VN085549.D
 Acq On : 29 Jan 2025 11:42
 Operator : JC\MD
 Sample : VN0129MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0129MBL01

Quant Time: Jan 30 00:30:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185591	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	345014	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	300844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	170931	57.056	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.120%			
35) Dibromofluoromethane	8.165	113	127945	53.454	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.900%			
50) Toluene-d8	10.565	98	421114	49.518	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.040%			
62) 4-Bromofluorobenzene	12.847	95	128268	44.092	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 88.180%			

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
Data File : VN085549.D
Acq On : 29 Jan 2025 11:42
Operator : JC\MD
Sample : VN0129MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0129MBL01

Quant Time: Jan 30 00:30:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

