

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
 Data File : VN085554.D
 Acq On : 29 Jan 2025 13:52
 Operator : JC\MD
 Sample : Q1205-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-236

Quant Time: Jan 30 00:34:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	172063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	326014	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	306107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	108073	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	156416	56.316	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 112.640%		
35) Dibromofluoromethane	8.165	113	117397	51.906	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.820%		
50) Toluene-d8	10.565	98	424827	52.866	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 105.740%		
62) 4-Bromofluorobenzene	12.847	95	121591	44.233	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 88.460%		
Target Compounds						
						Qvalue
16) Acetone	4.424	43	77976	86.889	ug/l	95
25) 2-Butanone	7.488	43	7968	6.033	ug/l	97
43) Isopropyl Acetate	8.688	43	116481	22.688	ug/l #	87
95) Naphthalene	15.641	128	16913	3.484	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
Data File : VN085554.D
Acq On : 29 Jan 2025 13:52
Operator : JC\MD
Sample : Q1205-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-236

Quant Time: Jan 30 00:34:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

