

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
 Data File : VN085559.D
 Acq On : 29 Jan 2025 15:53
 Operator : JC\MD
 Sample : Q1207-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-4.5-012725

Quant Time: Jan 30 00:37:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	160119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302819	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	292646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	106369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	148132	57.312	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.620%	
35) Dibromofluoromethane	8.165	113	111026	52.849	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	105.700%	
50) Toluene-d8	10.565	98	396088	53.065	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	106.140%	
62) 4-Bromofluorobenzene	12.847	95	119098	46.645	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	93.280%	
Target Compounds						
16) Acetone	4.430	43	73891	88.479	ug/l	99
25) 2-Butanone	7.489	43	9272	7.545	ug/l #	86
29) Tetrahydrofuran	7.847	42	3958	5.080	ug/l	95
43) Isopropyl Acetate	8.688	43	23346	4.896	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
Data File : VN085559.D
Acq On : 29 Jan 2025 15:53
Operator : JC\MD
Sample : Q1207-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-4.5-012725

Quant Time: Jan 30 00:37:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

