

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
 Data File : VN085561.D
 Acq On : 29 Jan 2025 16:41
 Operator : JC\MD
 Sample : Q1207-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JPP-20.2-012725

Quant Time: Jan 30 00:38:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	159671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	312230	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298965	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	104378	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150154	58.257	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 116.520%		
35) Dibromofluoromethane	8.165	113	111970	51.692	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.380%		
50) Toluene-d8	10.565	98	411896	53.520	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.040%		
62) 4-Bromofluorobenzene	12.847	95	116737	44.342	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 88.680%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	67993	81.645	ug/l	98
25) 2-Butanone	7.488	43	6742	5.501	ug/l	92
43) Isopropyl Acetate	8.688	43	115617	23.514	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
Data File : VN085561.D
Acq On : 29 Jan 2025 16:41
Operator : JC\MD
Sample : Q1207-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JPP-20.2-012725

Quant Time: Jan 30 00:38:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

