

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
 Data File : VN085562.D
 Acq On : 29 Jan 2025 17:05
 Operator : JC\MD
 Sample : Q1209-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-4

Quant Time: Jan 30 00:38:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	321449	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	308281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	107199	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	152478	57.381	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 114.760%		
35) Dibromofluoromethane	8.165	113	113204	50.763	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.520%		
50) Toluene-d8	10.565	98	422385	53.309	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.620%		
62) 4-Bromofluorobenzene	12.847	95	121611	44.868	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 89.740%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	73157	85.205	ug/l	93
25) 2-Butanone	7.489	43	7124	5.638	ug/l	100
43) Isopropyl Acetate	8.688	43	125912	24.874	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
Data File : VN085562.D
Acq On : 29 Jan 2025 17:05
Operator : JC\MD
Sample : Q1209-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-4

Quant Time: Jan 30 00:38:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

