

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
 Data File : VN085570.D
 Acq On : 29 Jan 2025 20:19
 Operator : JC\MD
 Sample : PB166319ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166319ZHE#07

Quant Time: Jan 30 00:42:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	167177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	336311	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	318942	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	112024	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159950	59.272	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.540%			
35) Dibromofluoromethane	8.165	113	118527	50.801	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.600%			
50) Toluene-d8	10.565	98	439088	52.968	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.940%			
62) 4-Bromofluorobenzene	12.847	95	127448	44.944	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 89.880%			
Target Compounds						
16) Acetone	4.430	43	35372	40.567	ug/l	100
30) Chloroform	7.965	83	6555	1.610	ug/l	95
43) Isopropyl Acetate	8.688	43	226342	42.738	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
Data File : VN085570.D
Acq On : 29 Jan 2025 20:19
Operator : JC\MD
Sample : PB166319ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166319ZHE#07

Quant Time: Jan 30 00:42:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

