

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
 Data File : VN085573.D
 Acq On : 29 Jan 2025 21:31
 Operator : JC\MD
 Sample : PB166319ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166319ZHE#10

Quant Time: Jan 30 00:43:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	158804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	325032	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	303081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	107565	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150051	58.535	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	117.080%	
35) Dibromofluoromethane	8.165	113	113428	50.302	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	100.600%	
50) Toluene-d8	10.565	98	419323	52.339	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	104.680%	
62) 4-Bromofluorobenzene	12.847	95	122747	44.788	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	89.580%	
Target Compounds						
16) Acetone	4.430	43	65798	79.441	ug/l	98
25) 2-Butanone	7.489	43	6585	5.403	ug/l	100
43) Isopropyl Acetate	8.688	43	256586m	50.129	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012925\
Data File : VN085573.D
Acq On : 29 Jan 2025 21:31
Operator : JC\MD
Sample : PB166319ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166319ZHE#10

Quant Time: Jan 30 00:43:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

