

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013019\
 Data File : VN053561.D
 Acq On : 30 Jan 2019 10:12
 Operator : JC/SP
 Sample : VSTDICCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Jan 31 06:58:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 31 05:27:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	740828	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1116013	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	990104	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	481296	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	363782	48.95	ug/l	0.00
Spiked Amount			Recovery	=	97.90%	
35) Dibromofluoromethane	7.59	113	349022	49.49	ug/l	0.00
Spiked Amount			Recovery	=	98.98%	
50) Toluene-d8	10.09	98	1307421	49.38	ug/l	0.00
Spiked Amount			Recovery	=	98.76%	
62) 4-Bromofluorobenzene	12.40	95	437547	49.55	ug/l	0.00
Spiked Amount			Recovery	=	99.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.85	85	415066	49.699	ug/l	99
3) Chloromethane	2.06	50	438756	49.886	ug/l	99
4) Vinyl Chloride	2.18	62	430182	50.386	ug/l	100
5) Bromomethane	2.57	94	272839	48.875	ug/l	98
6) Chloroethane	2.71	64	240502	49.294	ug/l	99
7) Trichlorofluoromethane	3.02	101	573742	49.555	ug/l	100
8) Diethyl Ether	3.41	74	209026	49.435	ug/l	98
9) 1,1,2-Trichlorotrifluoroet	3.76	101	349505	48.811	ug/l	98
10) Methyl Iodide	3.95	142	555572	49.434	ug/l	99
11) Tert butyl alcohol	4.78	59	107063	243.413	ug/l	99
12) 1,1-Dichloroethene	3.74	96	340936	49.697	ug/l	97
13) Acrolein	3.61	56	47714	251.781	ug/l	96
14) Allyl chloride	4.33	41	547332	50.108	ug/l	99
15) Acrylonitrile	4.99	53	596779	251.331	ug/l	100
16) Acetone	3.81	43	443233	232.508	ug/l	99
17) Carbon Disulfide	4.06	76	1050034	49.156	ug/l	99
18) Methyl Acetate	4.32	43	267378	47.361	ug/l	98
19) Methyl tert-butyl Ether	5.05	73	885196	50.900	ug/l	100
20) Methylene Chloride	4.55	84	376196	47.038	ug/l	98
21) trans-1,2-Dichloroethene	5.05	96	364414	49.237	ug/l	99
22) Diisopropyl ether	5.95	45	1031863	50.983	ug/l	96
23) Vinyl Acetate	5.90	43	3658025	261.263	ug/l	99
24) 1,1-Dichloroethane	5.85	63	660616	48.936	ug/l	99
25) 2-Butanone	6.83	43	676628	248.869	ug/l	99
26) 2,2-Dichloropropane	6.82	77	506270	49.129	ug/l	100
27) cis-1,2-Dichloroethene	6.83	96	411865	50.045	ug/l	98
28) Bromochloromethane	7.20	49	289010	49.334	ug/l	97
29) Tetrahydrofuran	7.21	42	445663	255.272	ug/l	98
30) Chloroform	7.37	83	642866	48.040	ug/l	98
31) Cyclohexane	7.65	56	613687	48.823	ug/l	98
32) 1,1,1-Trichloroethane	7.57	97	550347	49.418	ug/l	99
36) 1,1-Dichloropropene	7.79	75	498417	49.544	ug/l	98
37) Ethyl Acetate	6.93	43	300551	53.492	ug/l	100
38) Carbon Tetrachloride	7.77	117	497399	49.080	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013019\
 Data File : VN053561.D
 Acq On : 30 Jan 2019 10:12
 Operator : JC/SP
 Sample : VSTDICCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Jan 31 06:58:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 31 05:27:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.08	83	611351	51.222	ug/l	98
40) Benzene	8.04	78	1528461	49.269	ug/l	100
41) Methacrylonitrile	7.17	41	155216	49.428	ug/l	91
42) 1,2-Dichloroethane	8.12	62	445742	48.390	ug/l	99
43) Isopropyl Acetate	8.17	43	520785	47.995	ug/l #	89
44) Trichloroethene	8.83	130	419833	47.535	ug/l	99
45) 1,2-Dichloropropane	9.12	63	395902	49.894	ug/l	100
46) Dibromomethane	9.21	93	231851	48.471	ug/l	97
47) Bromodichloromethane	9.40	83	487332	49.001	ug/l	99
48) Methyl methacrylate	9.20	41	254695	49.173	ug/l	95
49) 1,4-Dioxane	9.20	88	69788	932.609	ug/l	96
51) 4-Methyl-2-Pentanone	9.99	43	1378287	258.488	ug/l	99
52) Toluene	10.16	92	923587	50.162	ug/l	100
53) t-1,3-Dichloropropene	10.38	75	501393	50.347	ug/l	99
54) cis-1,3-Dichloropropene	9.84	75	592042	50.169	ug/l	99
55) 1,1,2-Trichloroethane	10.56	97	332636	50.137	ug/l	97
56) Ethyl methacrylate	10.43	69	408788	52.083	ug/l	99
57) 1,3-Dichloropropane	10.71	76	555882	49.024	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.69	63	1166913	240.061	ug/l	99
59) 2-Hexanone	10.75	43	921439	254.554	ug/l	99
60) Dibromochloromethane	10.90	129	376106	50.270	ug/l	99
61) 1,2-Dibromoethane	11.01	107	321027	48.887	ug/l	100
64) Tetrachloroethene	10.63	164	411447	48.478	ug/l	99
65) Chlorobenzene	11.43	112	1015452	48.284	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.51	131	362254	49.321	ug/l	99
67) Ethyl Benzene	11.51	91	1714307	49.622	ug/l	100
68) m/p-Xylenes	11.62	106	1328400	100.748	ug/l	99
69) o-Xylene	11.95	106	644185	50.571	ug/l	99
70) Styrene	11.96	104	1051441	50.926	ug/l	100
71) Bromoform	12.13	173	256597	50.958	ug/l #	100
73) Isopropylbenzene	12.25	105	1660709	49.083	ug/l	100
74) N-amyl acetate	12.07	43	407569	50.094	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.50	83	384145	47.440	ug/l	100
76) 1,2,3-Trichloropropane	12.55	75	478393	77.385	ug/l #	100
77) Bromobenzene	12.53	156	444675	48.765	ug/l	98
78) n-propylbenzene	12.59	91	1940676	49.617	ug/l	100
79) 2-Chlorotoluene	12.68	91	1128391	48.627	ug/l	99
80) 1,3,5-Trimethylbenzene	12.73	105	1416222	50.477	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.30	75	111094	50.532	ug/l	100
82) 4-Chlorotoluene	12.77	91	1165893	48.963	ug/l	100
83) tert-Butylbenzene	12.99	119	1226009	49.441	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	1445967	50.585	ug/l	99
85) sec-Butylbenzene	13.17	105	1685900	49.961	ug/l	100
86) p-Isopropyltoluene	13.29	119	1493325	50.600	ug/l	100
87) 1,3-Dichlorobenzene	13.28	146	789011	48.133	ug/l	100
88) 1,4-Dichlorobenzene	13.36	146	781841	48.217	ug/l	100
89) n-Butylbenzene	13.62	91	1284871	50.299	ug/l	100
90) Hexachloroethane	13.88	117	235487	48.375	ug/l	99
91) 1,2-Dichlorobenzene	13.65	146	768253	48.955	ug/l	100
92) 1,2-Dibromo-3-Chloropropan	14.27	75	57315	49.665	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013019\
Data File : VN053561.D
Acq On : 30 Jan 2019 10:12
Operator : JC/SP
Sample : VSTDICCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Quant Time: Jan 31 06:58:15 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 31 05:27:12 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.91	180	477066	53.005	ug/l	99
94) Hexachlorobutadiene	15.01	225	276446	49.321	ug/l	100
95) Naphthalene	15.13	128	1033384	48.916	ug/l	99
96) 1,2,3-Trichlorobenzene	15.31	180	437456	52.382	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN013019\
 Data File : VN053561.D
 Acq On : 30 Jan 2019 10:12
 Operator : JC/SP
 Sample : VSTDICCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Quant Time: Jan 31 06:58:15 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 31 05:27:12 2019
 Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :


VSTDICCC050

Tot Ion:168 Resp: 740828

Ion Ratio Lower Upper




168 100

99 49.1 39.1 58.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#2

Dichlorodifluoromethane

Concen: 49.699 ug/l

RT: 1.85 min Scan# 24

Delta R.T. 0.00 min

Lab File: VN053561.D


Acq: 30 Jan 2019 10:12

Tgt Ion: 85 Resp: 415066

Ion Ratio Lower Upper



85 100

87 33.0 16.2 48.5



Abundance Ion 84.90 (84.60 to 85.60): VNO

Ion 86.90 (86.60 to 87.60): VNO









#3
 Chloromethane
 Concen: 49.886 ug/l
 RT: 2.06 min Scan# 88
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tot Ion: 50 Resp: 438756
 Ion Ratio Lower Upper
 50 100
 52 32.9 26.6 40.0



#4
 Vinyl Chloride
 Concen: 50.386 ug/l
 RT: 2.18 min Scan# 128
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion: 62 Resp: 430182
 Ion Ratio Lower Upper
 62 100
 64 32.0 25.4 38.2









#5
Bromomethane
Concen: 48.875 ug/l
RT: 2.57 min Scan# 248
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tot Ion: 94 Resp: 272839
Ion Ratio Lower Upper
94 100
96 94.1 76.8 115.2



#6
Chloroethane
Concen: 49.294 ug/l
RT: 2.71 min Scan# 290
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 64 Resp: 240502
Ion Ratio Lower Upper
64 100
66 32.2 25.1 37.7









#7

Trichlorofluoromethane
 Concen: 49.555 ug/l
 RT: 3.02 min Scan# 388
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050




Tot Ion:101 Resp: 573742
 Ion Ratio Lower Upper
 101 100
 103 64.2 51.6 77.4



#8

Diethyl Ether
 Concen: 49.435 ug/l
 RT: 3.41 min Scan# 509
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion: 74 Resp: 209026
 Ion Ratio Lower Upper
 74 100
 45 94.2 47.9 143.7





#9

1,1,2-Trichlorotrifluoroethane

Concen: 48.811 ug/l

RT: 3.76 min Scan# 619

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion:101 Resp: 349505

Ion Ratio Lower Upper

101 100

85 41.2 32.8 49.2

151 86.0 66.9 100.3






Abundance

Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): V

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 49.434 ug/l

RT: 3.95 min Scan# 678

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion:142 Resp: 555572

Ion Ratio Lower Upper

142 100

127 38.9 31.0 46.4

141 14.3 11.1 16.7






Abundance

Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V






#11




Tert butyl alcohol
Concen: 243.413 ug/l
RT: 4.78 min Scan# 935
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tot Ion: 59 Resp: 107063
Ion Ratio Lower Upper
59 100
57 10.0 8.2 12.4




Abundance Ion 59.00 (58.70 to 59.70): VN0
Ion 57.00 (56.70 to 57.70): VN0





#12


1,1-Dichloroethene
Concen: 49.697 ug/l
RT: 3.74 min Scan# 612
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 96 Resp: 340936
Ion Ratio Lower Upper
96 100
61 149.9 122.9 184.3
98 63.0 49.4 74.0



Abundance Ion 95.90 (95.60 to 96.60): VN0
Ion 60.90 (60.60 to 61.60): VN0
Ion 97.90 (97.60 to 98.60): VN0





#13

Acrolein

Concen: 251.781 ug/l

RT: 3.61 min Scan# 571

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :


VSTDICCC050

Tot Ion: 56 Resp: 47714

Ion Ratio Lower Upper




56 100

55 68.4 57.4 86.0



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 55.00 (54.70 to 55.70): VN0



#14

Allyl chloride

Concen: 50.108 ug/l

RT: 4.33 min Scan# 794

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion: 41 Resp: 547332

Ion Ratio Lower Upper

41 100

39 60.8 48.9 73.3



76 38.0 29.4 44.2



Abundance Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0

Ion 75.90 (75.60 to 76.60): VN0





#15

Acrylonitrile

Concen: 251.331 ug/l

RT: 4.99 min Scan# 999

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion: 53 Resp: 596779

Ion Ratio Lower Upper

53 100

52 82.5 66.0 99.0




51 35.4 28.3 42.5



Abundance Ion 53.00 (52.70 to 53.70): VN053561.D (-975) (-)

Ion 52.00 (51.70 to 52.70): VN053561.D (-975) (-)

Ion 51.00 (50.70 to 51.70): VN053561.D (-975) (-)



#16

Acetone

Concen: 232.508 ug/l

RT: 3.81 min Scan# 635

Delta R.T. 0.00 min

Lab File: VN053561.D


Acq: 30 Jan 2019 10:12

Tgt Ion: 43 Resp: 443233

Ion Ratio Lower Upper



43 100


58 34.5 27.1 40.7



Abundance Ion 43.00 (42.70 to 43.70): VN053561.D (-603) (-)

Ion 58.00 (57.70 to 58.70): VN053561.D (-603) (-)









#17
Carbon Disulfide
Concen: 49.156 ug/l
RT: 4.06 min Scan# 710
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tot Ion: 76 Resp: 1050034
Ion Ratio Lower Upper
76 100
78 9.4 7.3 10.9



#18
Methyl Acetate
Concen: 47.361 ug/l
RT: 4.32 min Scan# 793
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 43 Resp: 267378
Ion Ratio Lower Upper
43 100
74 25.3 19.4 29.2









#19
Methyl tert-butyl Ether
Concen: 50.900 ug/l
RT: 5.05 min Scan# 1018
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampled :
VSTDICCC050


Tot Ion: 73 Resp: 885196
Ion Ratio Lower Upper
73 100
57 22.8 18.3 27.5



#20
Methylene Chloride
Concen: 47.038 ug/l
RT: 4.55 min Scan# 864
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 84 Resp: 376196
Ion Ratio Lower Upper
84 100
49 114.0 94.3 141.5
51 36.1 29.4 44.2
86 64.9 51.4 77.0





#21

trans-1,2-Dichloroethene

Concen: 49.237 ug/l

RT: 5.05 min Scan# 1018

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion: 96 Resp: 364414

Ion Ratio Lower Upper

96 100

61 136.7 110.1 165.1

98 63.2 50.0 75.0






Abundance

Ion 95.90 (95.60 to 96.60): VN0

Ion 60.90 (60.60 to 61.60): VN0

Ion 97.90 (97.60 to 98.60): VN0



#22

Diisopropyl ether

Concen: 50.983 ug/l

RT: 5.95 min Scan# 1299

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Tgt Ion: 45 Resp: 1031863


Ion Ratio Lower Upper

45 100

43 60.8 44.2 66.4

87 28.4 22.7 34.1

59 11.9 9.4 14.2





Abundance


Ion 45.00 (44.70 to 45.70): VN0

Ion 43.00 (42.70 to 43.70): VN0

Ion 87.00 (86.70 to 87.70): VN0

Ion 59.00 (58.70 to 59.70): VN0





#23

Vinyl Acetate

Concen: 261.263 ug/l

RT: 5.90 min Scan# 1282

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12





Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tot Ion: 43 Resp: 3658025

Ion Ratio Lower Upper

43 100

86 10.8 8.4 12.6



#24

1,1-Dichloroethane

Concen: 48.936 ug/l

RT: 5.85 min Scan# 1268

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12




Tgt Ion: 63 Resp: 660616


Ion Ratio Lower Upper

63 100

98 7.1 3.3 9.9

100 4.4 2.1 6.5









#25
2-Butanone
Concen: 248.869 ug/l
RT: 6.83 min Scan# 1574
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampled :
VSTDICCC050

Tot Ion: 43 Resp: 676628
Ion Ratio Lower Upper
43 100
72 28.2 22.0 33.0



#26
2,2-Dichloropropane
Concen: 49.129 ug/l
RT: 6.82 min Scan# 1571
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 77 Resp: 506270
Ion Ratio Lower Upper
77 100
97 24.6 12.2 36.6





#27

cis-1,2-Dichloroethene

Concen: 50.045 ug/l

RT: 6.83 min Scan# 1573

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion: 96 Resp: 411865

Ion Ratio Lower Upper

96 100

61 139.1 0.0 283.8




98 64.4 0.0 129.2



Abundance Ion 95.90 (95.60 to 96.60): VN0

Ion 60.90 (60.60 to 61.60): VN0

Ion 97.90 (97.60 to 98.60): VN0



#28

Bromochloromethane

Concen: 49.334 ug/l

RT: 7.20 min Scan# 1687

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion: 49 Resp: 289010

Ion Ratio Lower Upper

49 100

129 2.1 0.0 4.0



130 86.0 66.3 99.5




Abundance Ion 48.90 (48.60 to 49.60): VN0

Ion 128.80 (128.50 to 129.50): V

Ion 129.80 (129.50 to 130.50): V









#29
Tetrahydrofuran
Concen: 255.272 ug/l
RT: 7.21 min Scan# 1691
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tot Ion: 42 Resp: 445663
Ion Ratio Lower Upper
42 100
72 45.1 35.2 52.8
71 42.1 33.0 49.4




Abundance Ion 42.00 (41.70 to 42.70): VN0
Ion 72.00 (71.70 to 72.70): VN0
Ion 71.00 (70.70 to 71.70): VN0





#30
Chloroform
Concen: 48.040 ug/l
RT: 7.37 min Scan# 1741
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 83 Resp: 642866
Ion Ratio Lower Upper
83 100
85 66.8 52.2 78.2



Abundance Ion 82.90 (82.60 to 83.60): VN0
Ion 84.90 (84.60 to 85.60): VN0





#31

Cyclohexane

Concen: 48.823 ug/l

RT: 7.65 min Scan# 1829

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion: 56 Resp: 613687

Ion Ratio Lower Upper




56 100

69 32.7 26.4 39.6

84 87.1 67.9 101.9



Abundance Ion 56.00 (55.70 to 56.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 84.00 (83.70 to 84.70): VNO



#32

1,1,1-Trichloroethane

Concen: 49.418 ug/l

RT: 7.57 min Scan# 1803

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion: 97 Resp: 550347

Ion Ratio Lower Upper



97 100


99 64.2 52.0 78.0

61 42.9 34.4 51.6



Abundance Ion 96.90 (96.60 to 97.60): VNO
Ion 98.90 (98.60 to 99.60): VNO
Ion 60.90 (60.60 to 61.60): VNO









#33
 1,2-Dichloroethane-d4
 Concen: 49.418 ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Tot Ion: 65 Resp: 363782
 Ion Ratio Lower Upper
 65 100
 67 55.3 0.0 111.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion: 114 Resp: 1116013
 Ion Ratio Lower Upper
 114 100
 63 19.4 0.0 38.6
 88 15.3 0.0 30.4









#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Tot Ion:113 Resp: 349022
 Ion Ratio Lower Upper
 113 100
 111 102.1 81.1 121.7
 192 23.5 18.2 27.2




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#36
 1,1-Dichloropropene
 Concen: 49.544 ug/l
 RT: 7.79 min Scan# 1872
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion: 75 Resp: 498417
 Ion Ratio Lower Upper
 75 100
 110 37.5 18.1 54.3
 77 32.0 25.1 37.7



Abundance Ion 74.90 (74.60 to 75.60): VNO
 Ion 109.90 (109.60 to 110.60): V
 Ion 76.90 (76.60 to 77.60): VNO









#37
Ethyl Acetate
Concen: 53.492 ug/l
RT: 6.93 min Scan# 1603
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tot Ion: 43 Resp: 300551
Ion Ratio Lower Upper
43 100
61 14.8 11.8 17.6
70 10.8 8.6 12.8




Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 60.90 (60.60 to 61.60): VN0
Ion 70.00 (69.70 to 70.70): VN0





#38
Carbon Tetrachloride
Concen: 49.080 ug/l
RT: 7.77 min Scan# 1866
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 117 Resp: 497399
Ion Ratio Lower Upper
117 100
119 94.8 77.2 115.8
121 29.6 24.6 37.0



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V









#39
Methvlcvclohexane
Concen: 51.222 ug/l
RT: 9.08 min Scan# 2273
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tot Ion: 83 Resp: 611351
Ion Ratio Lower Upper
83 100
55 72.5 60.1 90.1
98 46.3 36.6 55.0



#40
Benzene
Concen: 49.269 ug/l
RT: 8.04 min Scan# 1950
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 78 Resp: 1528461
Ion Ratio Lower Upper
78 100
77 23.3 18.8 28.2









#41
Methacrylonitrile
Concen: 49.428 ug/l
RT: 7.17 min Scan# 1678
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Instrument :
MSVOA_N
ClientSampled :
VSTDICCC050

Tot Ion: 41 Resp: 155216
Ion Ratio Lower Upper
41 100
39 59.7 46.8 70.2
67 91.5 63.0 94.6
52 35.1 23.8 35.8




Abundance Ion 41.00 (40.70 to 41.70): VNC
Ion 39.00 (38.70 to 39.70): VNC
Ion 67.00 (66.70 to 67.70): VNC
Ion 52.00 (51.70 to 52.70): VNC






#42
1,2-Dichloroethane
Concen: 48.390 ug/l
RT: 8.12 min Scan# 1975
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 62 Resp: 445742
Ion Ratio Lower Upper
62 100
98 10.0 0.0 19.6



Abundance Ion 61.90 (61.60 to 62.60): VNC
Ion 97.90 (97.60 to 98.60): VNC





#43

Isopropyl Acetate

Concen: 47.995 ug/l

RT: 8.17 min Scan# 1988

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion: 43 Resp: 520785

Ion Ratio Lower Upper

43 100

61 20.9 23.9 35.9#




87 13.8 11.0 16.6



Abundance Ion 43.00 (42.70 to 43.70): VN053561.D (-1968) (-)

Ion 61.00 (60.70 to 61.70): VN053561.D (-1968) (-)

Ion 87.00 (86.70 to 87.70): VN053561.D (-1968) (-)



#44

Trichloroethene

Concen: 47.535 ug/l

RT: 8.83 min Scan# 2196

Delta R.T. 0.00 min

Lab File: VN053561.D


Acq: 30 Jan 2019 10:12

Tgt Ion:130 Resp: 419833

Ion Ratio Lower Upper



130 100


95 94.3 0.0 191.4



Abundance Ion 129.80 (129.50 to 130.50): VN053561.D (-2179) (-)

Ion 94.90 (94.60 to 95.60): VN053561.D (-2179) (-)





#45

1,2-Dichloropropane

Concen: 49.894 ug/l

RT: 9.12 min Scan# 2285

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :


VSTDICCC050

Tot Ion: 63 Resp: 395902

Ion Ratio Lower Upper

63 100

65 31.1 25.0 37.4



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 64.90 (64.60 to 65.60): VN0

9.12

200000



150000

100000

50000

0

Time-->



#46

Dibromomethane

Concen: 48.471 ug/l

RT: 9.21 min Scan# 2312

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion: 93 Resp: 231851

Ion Ratio Lower Upper

93 100

95 84.3 67.3 100.9

174 119.2 91.0 136.4



Abundance Ion 92.80 (92.50 to 93.50): VN0

Ion 94.80 (94.50 to 95.50): VN0

Ion 173.70 (173.40 to 174.40): V


150000

100000

50000

0

Time-->





#47

Bromodichloromethane

Concen: 49.001 ug/l

RT: 9.40 min Scan# 2373

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion: 83 Resp: 487332

Ion Ratio Lower Upper

83 100

85 64.8 51.0 76.6

127 9.1 7.0 10.6






Abundance Ion 82.90 (82.60 to 83.60): VN053561.D (-2355) (-)

Ion 84.90 (84.60 to 85.60): VN053561.D (-2355) (-)

Ion 126.80 (126.50 to 127.50): VN053561.D (-2355) (-)

Time-->



#48

Methyl methacrylate

Concen: 49.173 ug/l

RT: 9.20 min Scan# 2309

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion: 41 Resp: 254695

Ion Ratio Lower Upper

41 100

69 97.5 73.8 110.6

39 53.5 40.0 60.0






Abundance Ion 41.00 (40.70 to 41.70): VN053561.D (-2309) (-)

Ion 69.00 (68.70 to 69.70): VN053561.D (-2309) (-)

Ion 39.00 (38.70 to 39.70): VN053561.D (-2309) (-)

Time-->









#49
1,4-Dioxane
Concen: 932.609 ug/l
RT: 9.20 min Scan# 2310
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tot Ion: 88 Resp: 69788
Ion Ratio Lower Upper
88 100
43 30.0 27.8 41.6
58 68.6 56.5 84.7




Abundance Ion 88.00 (87.70 to 88.70): VN0
Ion 43.00 (42.70 to 43.70): VN0
Ion 58.00 (57.70 to 58.70): VN0






#50
Toluene-d8
Concen: 932.609 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 98 Resp: 1307421
Ion Ratio Lower Upper
98 100
100 64.8 51.4 77.0



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 258.488 ug/l

RT: 9.99 min Scan# 2554

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tot Ion: 43 Resp: 1378287

Ion Ratio Lower Upper

43 100

58 40.7 32.8 49.2



#52

Toluene

Concen: 50.162 ug/l

RT: 10.16 min Scan# 2607

Delta R.T. 0.00 min

Lab File: VN053561.D




Acq: 30 Jan 2019 10:12


Tgt Ion: 92 Resp: 923587

Ion Ratio Lower Upper

92 100

91 174.9 139.9 209.9









#53
t-1,3-Dichloropropene
Concen: 50.347 ug/l
RT: 10.38 min Scan# 2677
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tot Ion: 75 Resp: 501393
Ion Ratio Lower Upper
75 100
77 31.3 25.5 38.3



#54
cis-1,3-Dichloropropene
Concen: 50.169 ug/l
RT: 9.84 min Scan# 2508
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 75 Resp: 592042
Ion Ratio Lower Upper
75 100
77 31.7 25.8 38.6
39 40.8 33.4 50.0









#55
 1,1,2-Trichloroethane
 Concen: 50.137 ug/l
 RT: 10.56 min Scan# 2733
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tot Ion	97	Resp	332636
Ion	Ratio	Lower	Upper
97	100		
83	83.9	70.1	105.1
85	55.8	45.4	68.0
99	60.1	49.8	74.8



#56
 Ethyl methacrylate
 Concen: 52.083 ug/l
 RT: 10.43 min Scan# 2692
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion	69	Resp	408788
Ion	Ratio	Lower	Upper
69	100		
41	60.9	49.4	74.0
39	28.7	23.3	34.9









#57
 1,3-Dichloropropane
 Concen: 49.024 ug/l
 RT: 10.71 min Scan# 2779
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tot Ion: 76 Resp: 555882
 Ion Ratio Lower Upper
 76 100
 78 32.6 25.6 38.4



#58
 2-Chloroethyl Vinyl ether
 Concen: 240.061 ug/l
 RT: 9.69 min Scan# 2463
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion: 63 Resp: 1166913
 Ion Ratio Lower Upper
 63 100
 106 27.4 21.5 32.3









#59
2-Hexanone
Concen: 254.554 ug/l
RT: 10.75 min Scan# 2793
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tot Ion: 43 Resp: 921439
Ion Ratio Lower Upper
43 100
58 57.7 28.6 85.8



#60
Dibromochloromethane
Concen: 50.270 ug/l
RT: 10.90 min Scan# 2839
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 129 Resp: 376106
Ion Ratio Lower Upper
129 100
127 76.4 38.6 115.8









#61
 1,2-Dibromoethane
 Concen: 48.887 ug/l
 RT: 11.01 min Scan# 2872
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050


Tot Ion:107 Resp: 321027
 Ion Ratio Lower Upper
 107 100
 109 94.2 75.7 113.5



#62
 4-Bromofluorobenzene
 Concen: 48.887 ug/l
 RT: 12.40 min Scan# 3306
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion: 95 Resp: 437547
 Ion Ratio Lower Upper
 95 100
 174 92.7 0.0 183.2
 176 90.6 0.0 177.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tot Ion:117 Resp: 990104
Ion Ratio Lower Upper
117 100
82 52.7 42.9 64.3
119 32.2 26.0 39.0



#64
Tetrachloroethene
Concen: 48.478 ug/l
RT: 10.63 min Scan# 2755
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion:164 Resp: 411447
Ion Ratio Lower Upper
164 100
166 127.2 102.2 153.4
129 89.3 72.6 109.0
131 85.9 69.5 104.3









#65
Chlorobenzene
Concen: 48.284 ug/l
RT: 11.43 min Scan# 3004
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Tot Ion:112 Resp: 1015452
Ion Ratio Lower Upper
112 100
114 31.9 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 49.321 ug/l
RT: 11.51 min Scan# 3028
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion:131 Resp: 362254
Ion Ratio Lower Upper
131 100
133 95.5 48.0 144.0
119 65.7 33.5 100.5









#67
Ethyl Benzene
Concen: 49.622 ug/l
RT: 11.51 min Scan# 3029
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampled :
VSTDICCC050

Tot Ion: 91 Resp: 1714307
Ion Ratio Lower Upper
91 100
106 31.3 25.0 37.6



#68
m/p-Xylenes
Concen: 100.748 ug/l
RT: 11.62 min Scan# 3063
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 106 Resp: 1328400
Ion Ratio Lower Upper
106 100
91 200.6 161.4 242.2









#69
o-Xylene
Concen: 50.571 ug/l
RT: 11.95 min Scan# 3164
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampled :
VSTDICCC050


Tot Ion:106 Resp: 644185
Ion Ratio Lower Upper
106 100
91 211.7 106.9 320.8



#70
Styrene
Concen: 50.926 ug/l
RT: 11.96 min Scan# 3169
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion:104 Resp: 1051441
Ion Ratio Lower Upper
104 100
78 47.6 38.2 57.2
103 55.9 44.4 66.6









#71
 Bromoform
 Concen: 50.958 ug/l
 RT: 12.13 min Scan# 3220
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Tot Ion:173 Resp: 256597
 Ion Ratio Lower Upper
 173 100
 175 49.5 24.6 74.0
 254 0.0 0.2 0.2#




Abundance Ion 172.70 (172.40 to 173.40): V
 Ion 174.70 (174.40 to 175.40): V
 Ion 254.40 (254.10 to 255.10): V






#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.35 min Scan# 3599
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion:152 Resp: 481296
 Ion Ratio Lower Upper
 152 100
 115 54.7 27.8 83.4
 150 172.6 0.0 349.0



Abundance Ion 151.90 (151.60 to 152.60): V
 Ion 114.90 (114.60 to 115.60): V
 Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 49.083 ug/l

RT: 12.25 min Scan# 3258

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tot Ion:105 Resp: 1660709

Ion Ratio Lower Upper

105 100

120 26.9 13.6 40.7



#74

N-ethyl acetate

Concen: 50.094 ug/l

RT: 12.07 min Scan# 3202

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Tgt Ion: 43 Resp: 407569




Ion Ratio Lower Upper


43 100

70 45.6 35.4 53.2

55 26.1 20.9 31.3

61 26.1 20.5 30.7





#75

1,1,2,2-Tetrachloroethane

Concen: 47.440 ug/l

RT: 12.50 min Scan# 3337

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tot Ion: 83 Resp: 384145

Ion Ratio Lower Upper

83 100

131 10.4 5.3 15.8

85 64.0 32.1 96.3



#76

1,2,3-Trichloropropane

Concen: 77.385 ug/l

RT: 12.55 min Scan# 3353

Delta R.T. 0.00 min

Lab File: VN053561.D




Acq: 30 Jan 2019 10:12


Tgt Ion: 75 Resp: 478393

Ion Ratio Lower Upper

75 100

77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 48.765 ug/l

RT: 12.53 min Scan# 3344

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion:156 Resp: 444675

Ion Ratio Lower Upper

156 100

77 171.3 87.9 263.8




158 97.8 48.8 146.4



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VN0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 49.617 ug/l

RT: 12.59 min Scan# 3364

Delta R.T. 0.00 min

Lab File: VN053561.D


Acq: 30 Jan 2019 10:12

Tgt Ion: 91 Resp: 1940676

Ion Ratio Lower Upper



91 100

120 23.6 11.9 35.5



Abundance Ion 91.00 (90.70 to 91.70): VN0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 48.627 ug/l

RT: 12.68 min Scan# 3391

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC050

Tot Ion: 91 Resp: 1128391

Ion Ratio Lower Upper

91 100

126 36.1 17.9 53.7



#80

1,3,5-Trimethylbenzene

Concen: 50.477 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. 0.00 min

Lab File: VN053561.D




Acq: 30 Jan 2019 10:12


Tgt Ion: 105 Resp: 1416222

Ion Ratio Lower Upper

105 100

120 49.8 24.9 74.6





#81

trans-1,4-Dichloro-2-butene

Concen: 50.532 ug/l

RT: 12.30 min Scan# 3274

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tot Ion: 75 Resp: 111094

Ion Ratio Lower Upper

75 100

53 87.1 69.8 104.6

89 44.5 35.6 53.4



#82

4-Chlorotoluene

Concen: 48.963 ug/l

RT: 12.77 min Scan# 3421

Delta R.T. 0.00 min

Lab File: VN053561.D




Acq: 30 Jan 2019 10:12


Tgt Ion: 91 Resp: 1165893

Ion Ratio Lower Upper

91 100

126 35.2 17.6 52.9









#83
 tert-Butylbenzene
 Concen: 49.441 ug/l
 RT: 12.99 min Scan# 3490
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Tot Ion:119 Resp: 1226009
 Ion Ratio Lower Upper
 119 100
 91 61.4 31.3 93.8
 134 26.4 12.9 38.7




Abundance Ion 119.00 (118.70 to 119.70): V
 Ion 91.00 (90.70 to 91.70): V
 Ion 134.00 (133.70 to 134.70): V






#84
 1,2,4-Trimethylbenzene
 Concen: 50.585 ug/l
 RT: 13.04 min Scan# 3504
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion:105 Resp: 1445967
 Ion Ratio Lower Upper
 105 100
 120 46.6 23.0 69.0



Abundance Ion 105.00 (104.70 to 105.70): V
 Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 49.961 ug/l

RT: 13.17 min Scan# 3545

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :


VSTDICCC050

Tot Ion:105 Resp: 1685900

Ion Ratio Lower Upper




105 100

134 20.4 10.2 30.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 50.600 ug/l

RT: 13.29 min Scan# 3582

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion:119 Resp: 1493325

Ion Ratio Lower Upper

119 100

134 26.9 13.5 40.4



91 21.7 11.0 33.0



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#87
 1,3-Dichlorobenzene
 Concen: 48.133 ug/l
 RT: 13.28 min Scan# 3580
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Tot Ion:146 Resp: 789011
 Ion Ratio Lower Upper
 146 100
 111 37.1 18.8 56.3
 148 64.4 32.0 96.2



#88
 1,4-Dichlorobenzene
 Concen: 48.217 ug/l
 RT: 13.36 min Scan# 3605
 Delta R.T. 0.00 min
 Lab File: VN053561.D
 Acq: 30 Jan 2019 10:12

Tgt Ion:146 Resp: 781841
 Ion Ratio Lower Upper
 146 100
 111 36.6 18.6 55.8
 148 64.7 32.4 97.0









#89
n-Butylbenzene
Concen: 50.299 ug/l
RT: 13.62 min Scan# 3683
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12




Instrument :
MSVOA_N
ClientSampled :
VSTDICCC050


Tot Ion: 91 Resp: 1284871
Ion Ratio Lower Upper
91 100
92 52.2 26.3 78.8
134 27.1 13.5 40.5



#90
Hexachloroethane
Concen: 48.375 ug/l
RT: 13.88 min Scan# 3764
Delta R.T. 0.00 min
Lab File: VN053561.D
Acq: 30 Jan 2019 10:12

Tgt Ion: 117 Resp: 235487
Ion Ratio Lower Upper
117 100
201 94.8 46.8 140.4





#91

1,2-Dichlorobenzene

Concen: 48.955 ug/l

RT: 13.65 min Scan# 3695

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion:146 Resp: 768253

Ion Ratio Lower Upper

146 100

111 38.7 19.5 58.5




148 64.2 31.9 95.9



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 49.665 ug/l

RT: 14.27 min Scan# 3887

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion: 75 Resp: 57315

Ion Ratio Lower Upper

75 100

155 100.3 48.9 146.7



157 127.7 63.7 191.3



Abundance Ion 74.90 (74.60 to 75.60): V

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V





#93

1,2,4-Trichlorobenzene

Concen: 53.005 ug/l

RT: 14.91 min Scan# 4085

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050


Tot Ion:180 Resp: 477066

Ion Ratio Lower Upper




180 100

182 95.7 48.1 144.3

145 27.2 13.9 41.6



Abundance Ion 179.80 (179.50 to 180.50): V
Ion 181.80 (181.50 to 182.50): V
Ion 144.80 (144.50 to 145.50): V



#94

Hexachlorobutadiene

Concen: 49.321 ug/l

RT: 15.01 min Scan# 4117

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12


Tgt Ion:225 Resp: 276446

Ion Ratio Lower Upper



225 100

223 62.0 31.1 93.2

227 63.5 32.0 96.2



Abundance Ion 224.70 (224.40 to 225.40): V
Ion 222.70 (222.40 to 223.40): V
Ion 226.70 (226.40 to 227.40): V





#95

Naphthalene

Concen: 48.916 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Instrument :


MSVOA_N

ClientSampleId :

VSTDICCC050

Tot Ion:128 Resp: 1033384




Ion	Ratio	Lower	Upper
128	100		
127	12.6	10.3	15.5
129	11.0	8.6	13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 52.382 ug/l

RT: 15.31 min Scan# 4211


Delta R.T. 0.00 min

Lab File: VN053561.D

Acq: 30 Jan 2019 10:12

Tgt Ion:180 Resp: 437456

Ion	Ratio	Lower	Upper
180	100		
182	96.7	47.9	143.8
145	30.2	14.7	44.1



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

