

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN013019\
 Data File : VN053567.D
 Acq On : 30 Jan 2019 13:14
 Operator : JC/SP
 Sample : K1282-35
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TPW1

Quant Time: Jan 31 06:25:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 31 05:27:12 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	811358	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1311728	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1160552	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	494573	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	423647	52.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.10%	
35) Dibromofluoromethane	7.59	113	418358	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
50) Toluene-d8	10.09	98	1553201	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
62) 4-Bromofluorobenzene	12.40	95	511500	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
Target Compounds						
16) Acetone	3.82	43	8267	3.960	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN013019\
Data File : VN053567.D
Acq On : 30 Jan 2019 13:14
Operator : JC/SP
Sample : K1282-35
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TPW1

Quant Time: Jan 31 06:25:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 31 05:27:12 2019
Response via : Initial Calibration

