

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN013019\
 Data File : VN053571.D
 Acq On : 30 Jan 2019 19:32
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 06:32:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 31 05:27:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.515	0.582	-13.0	94	0.00
3 P	Chloromethane	0.594	0.623	-4.9	95	0.00
4 C	Vinyl Chloride	0.576	0.617	-7.1#	96	0.00
5 T	Bromomethane	0.377	0.387	-2.7	95	0.00
6 T	Chloroethane	0.329	0.347	-5.5	97	0.00
7 T	Trichlorofluoromethane	0.781	0.810	-3.7	95	0.00
8 T	Diethyl Ether	0.285	0.297	-4.2	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.483	0.491	-1.7	94	0.00
10 T	Methyl Iodide	0.759	0.779	-2.6	94	0.00
11 T	Tert butyl alcohol	0.030	0.032	-6.7	99	0.00
12 CM	1,1-Dichloroethene	0.463	0.485	-4.8#	95	0.00
13 T	Acrolein	0.016	0.023	-43.7#	160#	0.00
14 T	Allyl chloride	0.737	0.766	-3.9	94	0.00
15 T	Acrylonitrile	0.160	0.177	-10.6	99	0.00
16 T	Acetone	0.129	0.116	10.1	88	0.00
17 T	Carbon Disulfide	1.442	1.461	-1.3	93	0.00
18 T	Methyl Acetate	0.381	0.386	-1.3	97	0.00
19 T	Methyl tert-butyl Ether	1.174	1.272	-8.3	96	0.00
20 T	Methylene Chloride	0.540	0.546	-1.1	97	0.00
21 T	trans-1,2-Dichloroethene	0.500	0.514	-2.8	95	0.00
22 T	Diisopropyl ether	1.366	1.464	-7.2	95	0.00
23 T	Vinyl Acetate	0.945	1.052	-11.3	96	0.00
24 P	1,1-Dichloroethane	0.911	0.944	-3.6	96	0.00
25 T	2-Butanone	0.183	0.189	-3.3	94	0.00
26 T	2,2-Dichloropropane	0.696	0.659	5.3	87	0.00
27 T	cis-1,2-Dichloroethene	0.555	0.582	-4.9	95	0.00
28 T	Bromochloromethane	0.395	0.413	-4.6	96	0.00
29 T	Tetrahydrofuran	0.118	0.131	-11.0	99	0.00
30 C	Chloroform	0.903	0.930	-3.0#	97	0.00
31 T	Cyclohexane	0.949	0.862	9.2	94	0.00
32 T	1,1,1-Trichloroethane	0.752	0.788	-4.8	96	0.00
33 S	1,2-Dichloroethane-d4	0.502	0.530	-5.6	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.316	0.320	-1.3	97	0.00
36 T	1,1-Dichloropropene	0.451	0.452	-0.2	96	0.00
37 T	Ethyl Acetate	0.252	0.269	-6.7	95	0.00
38 T	Carbon Tetrachloride	0.454	0.448	1.3	95	0.00
39 T	Methylcyclohexane	0.535	0.527	1.5	91	0.00
40 TM	Benzene	1.390	1.379	0.8	96	0.00
41 T	Methacrylonitrile	0.141	0.141	0.0	96	0.00
42 TM	1,2-Dichloroethane	0.413	0.404	2.2	96	0.00
43 T	Isopropyl Acetate	0.486	0.490	-0.8	100	0.00
44 TM	Trichloroethene	0.396	0.377	4.8	95	0.00
45 C	1,2-Dichloropropane	0.356	0.361	-1.4#	97	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN013019\
 Data File : VN053571.D
 Acq On : 30 Jan 2019 19:32
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 31 06:32:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 31 05:27:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.214	0.214	0.0	98	0.00
47 T	Bromodichloromethane	0.446	0.444	0.4	97	0.00
48 T	Methyl methacrylate	0.232	0.229	1.3	95	0.00
49 T	1,4-Dioxane	0.003	0.003	0.0	102	0.00
50 S	Toluene-d8	1.186	1.196	-0.8	97	0.00
51 T	4-Methyl-2-Pentanone	0.239	0.255	-6.7	98	0.00
52 CM	Toluene	0.825	0.838	-1.6#	96	0.00
53 T	t-1,3-Dichloropropene	0.446	0.449	-0.7	95	0.00
54 T	cis-1,3-Dichloropropene	0.529	0.533	-0.8	95	0.00
55 T	1,1,2-Trichloroethane	0.297	0.302	-1.7	96	0.00
56 T	Ethyl methacrylate	0.352	0.372	-5.7	97	0.00
57 T	1,3-Dichloropropane	0.508	0.514	-1.2	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.196	0.209	-6.6	95	0.00
59 T	2-Hexanone	0.162	0.163	-0.6	94	0.00
60 T	Dibromochloromethane	0.335	0.344	-2.7	97	0.00
61 T	1,2-Dibromoethane	0.294	0.293	0.3	97	0.00
62 S	4-Bromofluorobenzene	0.396	0.401	-1.3	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.429	0.419	2.3	95	0.00
65 PM	Chlorobenzene	1.062	1.040	2.1	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.375	-1.1	97	0.00
67 C	Ethyl Benzene	1.745	1.745	0.0	95	0.00
68 T	m/p-Xylenes	0.666	0.673	-1.1	95	0.00
69 T	o-Xylene	0.643	0.658	-2.3	96	0.00
70 T	Styrene	1.043	1.072	-2.8	95	0.00
71 P	Bromoform	0.254	0.270	-6.3	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.515	3.498	0.5	96	0.00
74 T	N-amyl acetate	0.845	0.855	-1.2	95	0.00
75 P	1,1,2,2-Tetrachloroethane	0.841	0.827	1.7	98	0.00
76 T	1,2,3-Trichloropropane	0.642	0.690	-7.5	106	0.00
77 T	Bromobenzene	0.947	0.940	0.7	96	0.00
78 T	n-propylbenzene	4.063	4.051	0.3	95	0.00
79 T	2-Chlorotoluene	2.411	2.400	0.5	97	0.00
80 T	1,3,5-Trimethylbenzene	2.915	3.007	-3.2	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.228	0.220	3.5	90	0.00
82 T	4-Chlorotoluene	2.474	2.461	0.5	96	0.00
83 T	tert-Butylbenzene	2.576	2.595	-0.7	96	0.00
84 T	1,2,4-Trimethylbenzene	2.970	3.074	-3.5	97	0.00
85 T	sec-Butylbenzene	3.506	3.534	-0.8	95	0.00
86 T	p-Isopropyltoluene	3.066	3.109	-1.4	95	0.00
87 T	1,3-Dichlorobenzene	1.703	1.662	2.4	96	0.00
88 T	1,4-Dichlorobenzene	1.685	1.633	3.1	95	0.00
89 T	n-Butylbenzene	2.654	2.649	0.2	94	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN013019\
Data File : VN053571.D
Acq On : 30 Jan 2019 19:32
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 31 06:32:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N013019W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 31 05:27:12 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.506	0.505	0.2	97	0.00
91 T	1,2-Dichlorobenzene	1.630	1.616	0.9	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.121	-0.8	96	0.00
93 T	1,2,4-Trichlorobenzene	0.935	0.982	-5.0	94	0.00
94 T	Hexachlorobutadiene	0.582	0.577	0.9	95	0.00
95 T	Naphthalene	1.905	2.156	-13.2	95	0.00
96 T	1,2,3-Trichlorobenzene	0.868	0.932	-7.4	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5