

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN013020\
 Data File : VN059931.D
 Acq On : 30 Jan 2020 16:08
 Operator : JC/MD
 Sample : L1014-15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-012920

Quant Time: Jan 30 17:15:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 18 03:09:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195715	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	297198	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	275947	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	123891	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	106863	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
35) Dibromofluoromethane	7.57	113	89985	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
50) Toluene-d8	10.08	98	357314	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
62) 4-Bromofluorobenzene	12.40	95	124496	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
Target Compounds						
43) Isopropyl Acetate	8.14	43	38547	9.232	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN013020\
Data File : VN059931.D
Acq On : 30 Jan 2020 16:08
Operator : JC/MD
Sample : L1014-15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-012920

Quant Time: Jan 30 17:15:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Sat Jan 18 03:09:32 2020
Response via : Initial Calibration

