

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
 Data File : VN076448.D
 Acq On : 30 Jan 2023 11:31
 Operator : JC\MD
 Sample : VN0130MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0130MBL01

Quant Time: Jan 31 00:34:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	361261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	590517	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	514474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	194147	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	221235	54.694	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.380%	
35) Dibromofluoromethane	8.171	113	184413	49.923	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.840%	
50) Toluene-d8	10.571	98	687260	50.378	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.760%	
62) 4-Bromofluorobenzene	12.853	95	211896	47.579	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.160%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
Data File : VN076448.D
Acq On : 30 Jan 2023 11:31
Operator : JC\MD
Sample : VN0130MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0130MBL01

Quant Time: Jan 31 00:34:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

