

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
 Data File : VN076452.D
 Acq On : 30 Jan 2023 13:17
 Operator : JC\MD
 Sample : PB150523TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150523TB

Quant Time: Jan 31 00:37:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	461486	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	768656	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	684660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	256396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	283945	54.952	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.900%
35) Dibromofluoromethane	8.177	113	236298	49.144	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.280%
50) Toluene-d8	10.571	98	904415	50.932	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.860%
62) 4-Bromofluorobenzene	12.853	95	287292	49.559	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.120%
Target Compounds						
						Qvalue
16) Acetone	4.453	43	44677	26.906	ug/l	98
20) Methylene Chloride	5.289	84	19988	3.518	ug/l	95
37) Ethyl Acetate	7.571	43	42535	7.792	ug/l	97
43) Isopropyl Acetate	8.694	43	215396	24.160	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
Data File : VN076452.D
Acq On : 30 Jan 2023 13:17
Operator : JC\MD
Sample : PB150523TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150523TB

Quant Time: Jan 31 00:37:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

