

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
 Data File : VN076461.D
 Acq On : 30 Jan 2023 16:53
 Operator : JC\MD
 Sample : PB150523ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150523ZHE#09

Quant Time: Jan 31 00:42:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	434287	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	738856	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	639823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	239811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	271379	55.809	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 111.620%		
35) Dibromofluoromethane	8.177	113	223658	48.391	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.780%		
50) Toluene-d8	10.571	98	862337	50.521	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.040%		
62) 4-Bromofluorobenzene	12.853	95	267754	48.051	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.100%		
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	42524	27.214	ug/l	100
20) Methylene Chloride	5.295	84	18960	3.551	ug/l	94
37) Ethyl Acetate	7.571	43	42753	8.147	ug/l	98
43) Isopropyl Acetate	8.694	43	458219	53.470	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
Data File : VN076461.D
Acq On : 30 Jan 2023 16:53
Operator : JC\MD
Sample : PB150523ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150523ZHE#09

Quant Time: Jan 31 00:42:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

