

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
 Data File : VN076465.D
 Acq On : 30 Jan 2023 18:29
 Operator : JC\MD
 Sample : 01329-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-6-BORING-2-7-14

Quant Time: Jan 31 00:44:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	459821	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	773829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	679469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	264973	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	279206	54.230	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.460%	
35) Dibromofluoromethane	8.177	113	235918	48.737	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.480%	
50) Toluene-d8	10.571	98	902523	50.485	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.980%	
62) 4-Bromofluorobenzene	12.853	95	282503	48.407	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.820%	
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	31552	19.049	ug/l	92
20) Methylene Chloride	5.289	84	37866	7.238	ug/l	96
37) Ethyl Acetate	7.571	43	37790	6.881	ug/l	98
43) Isopropyl Acetate	8.694	43	356865	39.761	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
Data File : VN076465.D
Acq On : 30 Jan 2023 18:29
Operator : JC\MD
Sample : 01329-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-6-BORING-2-7-14

Quant Time: Jan 31 00:44:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

