

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
 Data File : VN076467.D
 Acq On : 30 Jan 2023 19:18
 Operator : JC\MD
 Sample : 01329-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-8-BORING-2-21-28

Quant Time: Jan 31 00:45:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	428819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	736607	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	664381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	260598	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	268771	55.977	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.960%
35) Dibromofluoromethane	8.177	113	222840	48.362	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.720%
50) Toluene-d8	10.571	98	880924	51.767	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.540%
62) 4-Bromofluorobenzene	12.853	95	276890	49.843	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.680%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	37959	24.595	ug/l	92
20) Methylene Chloride	5.289	84	17433	3.265	ug/l	95
25) 2-Butanone	7.500	43	12632	5.373	ug/l	97
37) Ethyl Acetate	7.571	43	39887	7.626	ug/l	98
43) Isopropyl Acetate	8.694	43	394759	46.206	ug/l #	70
95) Naphthalene	15.641	128	74061	7.108	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
Data File : VN076467.D
Acq On : 30 Jan 2023 19:18
Operator : JC\MD
Sample : 01329-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-8-BORING-2-21-28

Quant Time: Jan 31 00:45:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

