

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
 Data File : VN076469.D
 Acq On : 30 Jan 2023 20:06
 Operator : JC\MD
 Sample : 01333-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-I

Quant Time: Jan 31 00:46:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	472081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	792959	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	711742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	268498	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	286729	54.245	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 108.480%		
35) Dibromofluoromethane	8.177	113	241593	48.705	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 97.420%		
50) Toluene-d8	10.571	98	928640	50.693	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 101.380%		
62) 4-Bromofluorobenzene	12.853	95	293292	49.043	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 98.080%		
Target Compounds						Qvalue
16) Acetone	4.447	43	35619	20.953	ug/l	92
20) Methylene Chloride	5.289	84	30863	5.621	ug/l	95
37) Ethyl Acetate	7.571	43	41394	7.353	ug/l	98
43) Isopropyl Acetate	8.694	43	449493	48.873	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN013023\
Data File : VN076469.D
Acq On : 30 Jan 2023 20:06
Operator : JC\MD
Sample : 01333-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-I

Quant Time: Jan 31 00:46:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

